



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118603

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/09/2022	16:05
2	CAL2	1	Water	NA	NA	20.2	7.00	02/09/2022	16:06
3	CAL3	1	Water	NA	NA	20.2	10.03	02/09/2022	16:07
4	ICV	1	Water	NA	NA	20.3	6.98	02/09/2022	16:09
5	CCV1	1	Water	NA	NA	20.4	2.01	02/09/2022	16:10
6	N1457-01	1	Solid	20.03	20	21.1	6.66	02/09/2022	16:15
7	N1457-01DUP	1	Solid	20.04	20	21.1	6.67	02/09/2022	16:16
8	CCV2	1	Water	NA	NA	20.5	12.03	02/09/2022	16:19

WORKLIST(Hardcopy Internal Chain)

118603

WorkList Name : phs n1459 WorkList ID : 156699 Department : Wet-Chemistry Date : 02-09-2022 15:56:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/09/2022	Solid	N1457-01	pH	Cool 4 deg C	PSEG03	M11	BUR-1140	02/09/2022	9045D

Date/Time 02/09/22 16:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 02/09/22 17:25
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)