



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118604

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	02/09/2022	16:05
2	CAL2	1	Water	NA	NA	20.2	7.00	02/09/2022	16:06
3	CAL3	1	Water	NA	NA	20.2	10.03	02/09/2022	16:07
4	ICV	1	Water	NA	NA	20.3	6.98	02/09/2022	16:09
5	CCV1	1	Water	NA	NA	20.7	2.02	02/09/2022	16:20
6	N1461-01	1	Water	NA	NA	20.2	10.48	02/09/2022	16:22
7	N1461-02	1	Water	NA	NA	20.3	8.32	02/09/2022	16:23
8	N1461-02DUP	1	Water	NA	NA	20.1	8.33	02/09/2022	16:25
9	CCV2	1	Water	NA	NA	20.1	12.03	02/09/2022	16:30

4B-118604

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w n1458 WorkList ID : 156698 Department : Wet-Chemistry Date : 02-09-2022 15:56:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/09/2022	Water	N1461-01	pH	Cool 4 deg C	PSEG03	M11	AUD-1253	02/09/2022	9040C
02/09/2022	Water	N1461-02	pH	Cool 4 deg C	PSEG03	M11	AUD-1254	02/09/2022	9040C

Date/Time 02/09/22 16:00
Raw Sample Received by: 7A(WC)
Raw Sample Relinquished by: 5P(SM)

Date/Time 02/04/22 17:25
Raw Sample Received by: 5P(SM)
Raw Sample Relinquished by: 7A(WC)