



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 3/1/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 02/10/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 02/11/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118612

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1451-01	YAWP8	1	1.14	8.39	9.53	7.65	77.6	
N1451-02	YAWP9	2	1.15	8.49	9.64	8.29	84.1	
N1451-03	YAWQ0	3	1.19	8.53	9.72	8.68	87.8	
N1451-04	YAWQ1	4	1.19	8.48	9.67	8.54	86.7	
N1451-05	YAWQ2	5	1.15	8.40	9.55	8.45	86.9	
N1451-06	YAWQ3	6	1.11	8.63	9.74	8.38	84.2	
N1451-07	YAWQ4	7	1.18	8.50	9.68	7.99	80.1	
N1451-08	YAX23	8	1.11	8.59	9.7	7.92	79.3	
N1451-09	YAX23D	9	1.11	8.59	9.7	7.92	79.3	
N1451-10	YAX23S	10	1.11	8.59	9.7	7.92	79.3	
N1451-11	YAWR5	11	1.16	8.61	9.77	9.36	95.2	
N1451-12	YAWR6	12	1.13	8.54	9.67	9.14	93.8	
N1451-13	YAWR7	13	1.11	8.52	9.63	9.08	93.5	
N1451-14	YAWR8	14	1.15	8.74	9.89	9.45	95.0	
N1451-15	YAWR9	15	1.12	8.79	9.91	9.43	94.5	
N1451-16	YAWS0	16	1.14	8.84	9.98	9.59	95.6	
N1451-17	YAWS1	17	1.15	8.62	9.77	9.36	95.2	
N1451-18	YAWS2	18	1.15	8.50	9.65	9.16	94.2	
N1451-19	YAWS3	19	1.19	8.56	9.75	8.28	82.8	
N1451-20	YAWS4	20	1.14	8.43	9.57	8.45	86.7	
N1451-21	YAWS5	21	1.12	8.56	9.68	8.58	87.1	
N1451-22	YAWS6	22	1.11	8.63	9.74	8.63	87.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$