



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/11/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:15  
In Date: 02/10/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:17  
Out Date: 02/11/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB118613

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| N1452-01 | YAWS7           | 1      | 1.16            | 8.60          | 9.76                     | 8.84                       | 89.3    |          |
| N1452-02 | YAWS7D          | 2      | 1.16            | 8.60          | 9.76                     | 8.84                       | 89.3    |          |
| N1452-03 | YAWS7S          | 3      | 1.16            | 8.60          | 9.76                     | 8.84                       | 89.3    |          |
| N1452-04 | YAWS8           | 4      | 1.19            | 8.54          | 9.73                     | 8.83                       | 89.5    |          |
| N1452-05 | YAWS9           | 5      | 1.19            | 8.58          | 9.77                     | 8.3                        | 82.9    |          |
| N1452-06 | YAWT0           | 6      | 1.16            | 8.40          | 9.56                     | 8.66                       | 89.3    |          |
| N1452-07 | YAWT2           | 7      | 1.13            | 8.70          | 9.83                     | 8.74                       | 87.5    |          |
| N1452-08 | YAWT3           | 8      | 1.16            | 8.55          | 9.71                     | 8.69                       | 88.1    |          |
| N1452-09 | YAWT4           | 9      | 1.18            | 8.76          | 9.94                     | 8.87                       | 87.8    |          |
| N1452-10 | YAWT5           | 10     | 1.17            | 8.50          | 9.67                     | 8.67                       | 88.2    |          |
| N1452-11 | YAWT6           | 11     | 1.19            | 8.63          | 9.82                     | 7.96                       | 78.4    |          |
| N1452-12 | YAWT7           | 12     | 1.18            | 8.59          | 9.77                     | 8.75                       | 88.1    |          |
| N1452-13 | YAWT8           | 13     | 1.19            | 8.53          | 9.72                     | 8.76                       | 88.7    |          |
| N1452-14 | YAWT9           | 14     | 1.19            | 8.71          | 9.9                      | 8.93                       | 88.9    |          |
| N1452-15 | YAWW2           | 15     | 1.18            | 8.48          | 9.66                     | 8.64                       | 88.0    |          |
| N1452-16 | YAWW3           | 16     | 1.15            | 8.48          | 9.63                     | 7.83                       | 78.8    |          |
| N1452-17 | YAWW4           | 17     | 1.18            | 8.75          | 9.93                     | 8.88                       | 88.0    |          |
| N1452-18 | YAWW5           | 18     | 1.18            | 8.45          | 9.63                     | 8.03                       | 81.1    |          |
| N1452-19 | YAWW7           | 19     | 1.16            | 8.45          | 9.61                     | 8.51                       | 87.0    |          |
| N1452-20 | YAWW8           | 20     | 1.12            | 8.76          | 9.88                     | 8.77                       | 87.3    |          |
| N1452-21 | YAWW9           | 21     | 1.18            | 8.80          | 9.98                     | 8.49                       | 83.1    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$