



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118616

Slope : 98.9

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2886         |
| buffer solution pH 7 yellow    | W2813         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2887         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.3             | 4.01        | 02/10/2022 | 13:40     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 02/10/2022 | 13:41     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.03       | 02/10/2022 | 13:42     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.4             | 6.98        | 02/10/2022 | 13:45     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.4             | 2.01        | 02/10/2022 | 13:46     |
| 6   | N1417-01    | 1  | Water  | NA          | NA          | 20.6             | 7.64        | 02/10/2022 | 13:48     |
| 7   | N1417-01DUP | 1  | Water  | NA          | NA          | 20.7             | 7.65        | 02/10/2022 | 13:49     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.4             | 12.03       | 02/10/2022 | 13:50     |

✓ 118616

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH W N1417      WorkList ID : 156724      Department : Wet-Chemistry      Date : 02-10-2022 13:30:48

| Due Date   | Matrix | Sample   | Test | Preservative | Customer     | Raw Sample Storage Location | Customer Sample   | Collect Date | Method |
|------------|--------|----------|------|--------------|--------------|-----------------------------|-------------------|--------------|--------|
| 02/03/2022 | Water  | N1417-01 | pH   |              | Cool 4 deg C | TULL02 N11                  | PIT-11-DEWATERING | 02/03/2022   | 9040C  |

Date/Time 02.10.22 13:35  
Raw Sample Received by: JN (WC)  
Raw Sample Relinquished by: JN (WC)

Date/Time 02.10.22 14:35  
Raw Sample Received by: JN (WC)  
Raw Sample Relinquished by: JN (WC)