



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118623

Slope : 99.1

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/10/2022	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	02/10/2022	15:36
3	CAL3	1	Water	NA	NA	20.4	10.03	02/10/2022	15:37
4	ICV	1	Water	NA	NA	20.4	7.01	02/10/2022	15:38
5	CCV1	1	Water	NA	NA	20.5	2.01	02/10/2022	15:40
6	N1444-02	1	Solid	20.03	20	20.1	9.75	02/10/2022	15:42
7	N1470-01	1	Solid	20.04	20	21.0	8.94	02/10/2022	15:45
8	N1491-01	1	Solid	20.03	20	20.7	7.15	02/10/2022	15:48
9	N1491-01DUP	1	Solid	20.04	20	20.6	7.16	02/10/2022	15:49
10	CCV2	1	Water	NA	NA	20.4	12.03	02/10/2022	15:50

VB 118623

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1470 WorkList ID : 156719 Department : Wet-Chemistry Date : 02-10-2022 12:36:29

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
02/09/2022	Solid	N1444-02	Corrosivity	Cool 4 deg C	PSEG03	M11	ERM-R-TPS	02/09/2022 9045D
02/09/2022	Solid	N1470-01	Corrosivity	Cool 4 deg C	PARS15	M11	CTY-WC-ROLLOFF	02/09/2022 9045D
02/10/2022	Solid	N1491-01	Corrosivity	Cool 4 deg C	PSEG03	M11	SOIL-DRUM	02/10/2022 9045D

Date/Time 02-10-22 15:30
Raw Sample Received by: SA (CSM)
Raw Sample Relinquished by: JD (CSM)

Date/Time 02-10-22 17:00
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: JD (CSM)