



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118624

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/10/2022	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	02/10/2022	15:36
3	CAL3	1	Water	NA	NA	20.4	10.03	02/10/2022	15:37
4	ICV	1	Water	NA	NA	20.4	7.01	02/10/2022	15:38
5	CCV1	1	Water	NA	NA	20.2	2.01	02/10/2022	16:45
6	N1491-02	1	Water	NA	NA	20.6	7.28	02/10/2022	16:46
7	N1491-03	1	Water	NA	NA	20.4	6.99	02/10/2022	16:47
8	N1491-03DUP	1	Water	NA	NA	20.5	7.00	02/10/2022	16:48
9	CCV2	1	Water	NA	NA	20.4	12.03	02/10/2022	16:49

WORKLIST(Hardcopy Internal Chain)

10118624

WorkList Name: CORROSIONITY N147W WorkList ID: 156732 Department: Wet-Chemistry Date: 02-10-2022 16:36:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/10/2022	Water	N1491-02	Corrosivity	Cool 4 deg C	PSEG03	M11	WATER-DRUM-1-3	02/10/2022	9040C
02/10/2022	Water	N1491-03	Corrosivity	Cool 4 deg C	PSEG03	M11	TOTE-17127	02/10/2022	9040C

Date/Time 02-10-22 16:40
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)

Date/Time 02-10-22 17:130
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)