



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118625

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/10/2022	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	02/10/2022	15:36
3	CAL3	1	Water	NA	NA	20.4	10.03	02/10/2022	15:37
4	ICV	1	Water	NA	NA	20.4	7.01	02/10/2022	15:38
5	CCV1	1	Water	NA	NA	20.2	2.02	02/10/2022	16:45
6	N1487-03	1	Water	NA	NA	20.1	6.46	02/10/2022	17:05
7	N1487-04	1	Water	NA	NA	20.1	6.56	02/10/2022	17:06
8	N1493-01	1	Water	NA	NA	20.2	5.45	02/10/2022	17:08
9	N1493-02	1	Water	NA	NA	20.2	5.16	02/10/2022	17:10
10	N1493-05	1	Water	NA	NA	20.2	6.72	02/10/2022	17:13
11	N1493-06	1	Water	NA	NA	20.3	6.82	02/10/2022	17:15
12	N1493-07	1	Water	NA	NA	20.3	6.74	02/10/2022	17:19
13	N1493-10	1	Water	NA	NA	20.3	6.06	02/10/2022	17:20
14	N1493-12	1	Water	NA	NA	20.6	5.68	02/10/2022	17:22
15	N1493-14	1	Water	NA	NA	20.6	6.92	02/10/2022	17:25
16	CCV2	1	Water	NA	NA	20.4	12.03	02/10/2022	17:26
17	N1493-16	1	Water	NA	NA	20.1	5.88	02/10/2022	17:29
18	N1493-16DUP	1	Water	NA	NA	20.1	5.89	02/10/2022	17:30
19	CCV3	1	Water	NA	NA	20.3	2.01	02/10/2022	17:32

VB118625

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1496 WorkList ID : 156735 Department : Wet-Chemistry Date : 02-10-2022 16:38:25

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
02/10/2022	Water	N1487-03	pH	Cool 4 deg C	PSEG03	M11 34046	02/10/2022	9040C
02/10/2022	Water	N1487-04	pH	Cool 4 deg C	PSEG03	M11 34047	02/10/2022	9040C
02/10/2022	Water	N1493-01	pH	Cool 4 deg C	PSEG03	M11 34050	02/10/2022	9040C
02/10/2022	Water	N1493-02	pH	Cool 4 deg C	PSEG03	M11 34051	02/10/2022	9040C
02/10/2022	Water	N1493-05	pH	Cool 4 deg C	PSEG03	M11 34052	02/10/2022	9040C
02/10/2022	Water	N1493-06	pH	Cool 4 deg C	PSEG03	M11 34053	02/10/2022	9040C
02/10/2022	Water	N1493-07	pH	Cool 4 deg C	PSEG03	M11 34054	02/10/2022	9040C
02/10/2022	Water	N1493-10	pH	Cool 4 deg C	PSEG03	M11 34049	02/10/2022	9040C
02/10/2022	Water	N1493-12	pH	Cool 4 deg C	PSEG03	M11 31370	02/10/2022	9040C
02/10/2022	Water	N1493-14	pH	Cool 4 deg C	PSEG03	M11 31371	02/10/2022	9040C
02/10/2022	Water	N1493-16	pH	Cool 4 deg C	PSEG03	M11 33576	02/10/2022	9040C

Date/Time : 02-10-22 16:40

Raw Sample Received by: JD CSM

Raw Sample Relinquished by: JD CSM

Date/Time : 02-10-22 18:00

Raw Sample Received by: JD CSM

Raw Sample Relinquished by: JD CSM