



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118629

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	02/11/2022	11:35
2	N1461-03		1	11	100.00	>212.0	>212.0	02/11/2022	11:55
3	N1461-03DUP		1	12	100.00	>212.0	>212.0	02/11/2022	12:20
4	N1487-05		1	15	100.00	>212.0	>212.0	02/11/2022	12:45
5	N1491-02		1	13	100.00	>212.0	>212.0	02/11/2022	13:40
6	N1491-03		1	14	100.00	>212.0	>212.0	02/11/2022	14:05
7	N1493-03		1	15	100.00	>212.0	>212.0	02/11/2022	14:32
8	N1493-08		1	14	100.00	>212.0	>212.0	02/11/2022	14:58
9	N1493-10		1	13	100.00	>212.0	>212.0	02/11/2022	15:15
10	N1493-12		1	16	100.00	>212.0	>212.0	02/11/2022	15:42
11	N1493-14		1	15	100.00	>212.0	>212.0	02/11/2022	16:05
12	N1493-16		1	14	100.00	>212.0	>212.0	02/11/2022	16:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118629

WorkList Name : FP21022 WorkList ID : 156743 Department : Wet-Chemistry Date : 02-10-2022 17:04:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2022	Water	N1461-03	Flash Point	Cool 4 deg C	PSEG03	M11	TOTE-COMP	02/09/2022	1010B
02/17/2022	Water	N1487-05	Flash Point	Cool 4 deg C	PSEG03	M11	34046-47	02/10/2022	1010B
02/17/2022	Water	N1491-02	Flash Point	Cool 4 deg C	PSEG03	M11	WATER-DRUM-1-3	02/10/2022	1010B
02/17/2022	Water	N1491-03	Flash Point	Cool 4 deg C	PSEG03	M11	TOTE-17127	02/10/2022	1010B
02/17/2022	Water	N1493-03	Flash Point	Cool 4 deg C	PSEG03	M11	34050-51	02/10/2022	1010B
02/17/2022	Water	N1493-08	Flash Point	Cool 4 deg C	PSEG03	M11	34052-53-54	02/10/2022	1010B
02/17/2022	Water	N1493-10	Flash Point	Cool 4 deg C	PSEG03	M11	34049	02/10/2022	1010B
02/17/2022	Water	N1493-12	Flash Point	Cool 4 deg C	PSEG03	M11	31370	02/10/2022	1010B
02/17/2022	Water	N1493-14	Flash Point	Cool 4 deg C	PSEG03	M11	31371	02/10/2022	1010B
02/17/2022	Water	N1493-16	Flash Point	Cool 4 deg C	PSEG03	M11	33576	02/10/2022	1010B

Date/Time 02/11/22 11:10
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)

Date/Time 02/11/22 16:30
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)