



# PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 2/14/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:00  
In Date: 02/11/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 02/12/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB118633

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| N1484-01 | JNLC0           | 1      | 1.14            | 8.42          | 9.56                     | 6.55                       | 64.3    |          |
| N1484-02 | JNLC1           | 2      | 1.14            | 8.41          | 9.55                     | 6.63                       | 65.3    |          |
| N1484-03 | JNLC2           | 3      | 1.17            | 8.51          | 9.68                     | 6.58                       | 63.6    |          |
| N1484-04 | JNLC3           | 4      | 1.15            | 7.35          | 8.5                      | 5.89                       | 64.5    |          |
| N1484-05 | JNLC4           | 5      | 1.12            | 8.51          | 9.63                     | 6.61                       | 64.5    |          |
| N1484-06 | JNLC5           | 6      | 1.15            | 8.80          | 9.95                     | 6.83                       | 64.5    |          |
| N1484-07 | JNLC6           | 7      | 1.17            | 8.13          | 9.3                      | 6.34                       | 63.6    |          |
| N1484-08 | JNLC7           | 8      | 1.17            | 8.80          | 9.97                     | 6.88                       | 64.9    |          |
| N1484-09 | JNLC8           | 9      | 1.13            | 8.27          | 9.4                      | 6.37                       | 63.4    |          |
| N1484-10 | JNLC8D          | 10     | 1.13            | 8.27          | 9.4                      | 6.37                       | 63.4    |          |
| N1484-11 | JNLC8S          | 11     | 1.13            | 8.27          | 9.4                      | 6.37                       | 63.4    |          |
| N1484-12 | JNLC9           | 12     | 1.15            | 8.24          | 9.39                     | 6.39                       | 63.6    |          |
| N1484-13 | JNLD0           | 13     | 1.16            | 8.38          | 9.54                     | 6.55                       | 64.3    |          |
| N1484-14 | JNLD1           | 14     | 1.15            | 8.44          | 9.59                     | 6.53                       | 63.7    |          |
| N1484-15 | JNLD2           | 15     | 1.18            | 8.69          | 9.87                     | 6.67                       | 63.2    |          |
| N1484-16 | JNLD3           | 16     | 1.11            | 8.73          | 9.84                     | 6.74                       | 64.5    |          |
| N1484-17 | JNLD4           | 17     | 1.17            | 8.51          | 9.68                     | 6.64                       | 64.3    |          |
| N1484-18 | JNLD5           | 18     | 1.19            | 8.75          | 9.94                     | 6.77                       | 63.8    |          |
| N1484-19 | JNLD6           | 19     | 1.13            | 8.38          | 9.51                     | 6.49                       | 64.0    |          |
| N1484-20 | JNLD7           | 20     | 1.18            | 8.46          | 9.64                     | 6.58                       | 63.8    |          |
| N1484-21 | JNLD8           | 21     | 1.11            | 6.21          | 7.32                     | 5.11                       | 64.4    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$