


LB118634

INSTRUMENT ID: CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	25	0.000	25				
0.2	0.200	0.184	-0.016	1046	0.0 %	1046				
2.5	2.500	2.472	-0.028	14079	0.0 %	14079				
5.0	5.000	5.017	0.017	28573	0.0 %	28573				
7.5	7.500	7.600	0.100	43282	0.0 %	43282				
10.0	10.000	9.922	-0.078	56505	0.0 %	56505				

LB118634 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Feb 2022 11:44:07

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50-0	25	-	0.00007471B	PPB	11 Feb 2022 13:05:25	S	Std
0.2 - 1	50-2	1046	-	0.20007471B	PPB	11 Feb 2022 13:07:42	S	Std
2.5 - 1	52-5	14079	-	2.50007471B	PPB	11 Feb 2022 13:09:59	S	Std
5.0 - 1	55-0	28573	-	5.00007471B	PPB	11 Feb 2022 13:12:16	S	Std
7.5 - 1	57-5	43282	-	7.50007471B	PPB	11 Feb 2022 13:14:33	S	Std
10.0 - 1	59-0	56505	-	10.00007471B	PPB	11 Feb 2022 13:16:50	S	Std
ICV89 - 1	ICV89	21967	3.8574	-7471B	PPB	11 Feb 2022 13:20:43	U	SMPL
ICB89 - 1	ICB89	-20	-0.0033	-7471B	PPB	11 Feb 2022 13:22:59	U	SMPL
CCV076 - 1	CCV076	28911	5.0767	-7471B	PPB	11 Feb 2022 13:25:13	U	SMPL
CCB076 - 1	CCB076	-28	-0.0047	-7471B	PPB	11 Feb 2022 13:27:29	U	SMPL
CRA - 1	CRA	1047	0.1841	-7471B	PPB	11 Feb 2022 13:29:44	U	SMPL
HighStd - 1	HighStd	56786	9.9712	-7471B	PPB	11 Feb 2022 13:31:59	U	SMPL
ChkStd - 1	ChkStd	39208	6.8847	-7471B	PPB	11 Feb 2022 13:34:15	U	SMPL
PB142593BL - 1	PBS	-41	-0.0070	-7471B	PPB	11 Feb 2022 13:36:32	U	SMPL
PB142593BS - 1	LCSS	21895	3.8447	-7471B	PPB	11 Feb 2022 13:38:49	U	SMPL
N1435-01 - 1	RO-349	1973	0.3467	-7471B	PPB	11 Feb 2022 13:41:06	U	SMPL
N1435-03 - 1	TP-369	919	0.1616	-7471B	PPB	11 Feb 2022 13:43:23	U	SMPL
N1436-02 - 1	289	5535	0.9721	-7471B	PPB	11 Feb 2022 13:45:40	U	SMPL
N1443-01 - 1	TP-1-SOUTH-COMP	5575	0.9791	-7471B	PPB	11 Feb 2022 13:47:57	U	SMPL
N1443-03 - 1	TP-1-NORTH-COMP	823	0.1447	-7471B	PPB	11 Feb 2022 13:50:12	U	SMPL
CCV077 - 1	CCV077	28868	5.0691	-7471B	PPB	11 Feb 2022 13:52:27	U	SMPL
CCB077 - 1	CCB077	-30	-0.0050	-7471B	PPB	11 Feb 2022 13:54:43	U	SMPL
N1444-02 - 1	ERMR-TPS	5551	0.9749	-7471B	PPB	11 Feb 2022 13:56:59	U	SMPL
N1449-01 - 1	20220040-NACOTE-2-COMPOSITE	4055	0.7122	-7471B	PPB	11 Feb 2022 13:59:15	U	SMPL
N1449-02 - 1	20220042-AMENDED-NACOTE-2-COMP	4233	0.7435	-7471B	PPB	11 Feb 2022 14:01:31	U	SMPL
N1455-01 - 1	TP-1	309159	54.2848	-7471B	PPB	11 Feb 2022 14:03:48	U	SMPL
N1455-03 - 1	TP-2	769797	HIGH	-7471B	PPB	11 Feb 2022 14:06:06	X	
N1455-05 - 1	TP-3	356301	HIGH	-7471B	PPB	11 Feb 2022 14:07:58	X	
N1455-07 - 1	TP-4	264176	46.3863	-7471B	PPB	11 Feb 2022 14:09:57	U	SMPL
N1455-09 - 1	TP-5	132563	23.2767	-7471B	PPB	11 Feb 2022 14:12:17	U	SMPL
N1457-01 - 1	BUR-1140	-20	-0.0033	-7471B	PPB	11 Feb 2022 14:14:37	U	SMPL
N1457-01DUP - 1	BUR-1140DUP	200	0.0353	-7471B	PPB	11 Feb 2022 14:16:55	U	SMPL
CCV078 - 1	CCV078	28597	5.0215	-7471B	PPB	11 Feb 2022 14:19:11	U	SMPL
CCB078 - 1	CCB078	-52	-0.0089	-7471B	PPB	11 Feb 2022 14:21:27	U	SMPL
N1457-01MS - 1	BUR-1140MS	23704	4.1624	-7471B	PPB	11 Feb 2022 14:30:34	U	SMPL
N1457-01MSD - 1	BUR-1140MSD	23018	4.0419	-7471B	PPB	11 Feb 2022 14:32:49	U	SMPL
N1469-01 - 1	B-1	6657	1.1691	-7471B	PPB	11 Feb 2022 14:35:05	U	SMPL
N1457-01LX5 - 1	BUR-1140	32	0.0059	-7471B	PPB	11 Feb 2022 14:37:21	U	SMPL
N1457-01A - 1	BUR-1140	22590	3.9668	-7471B	PPB	11 Feb 2022 14:39:38	U	SMPL
CCV079 - 1	CCV079	27586	4.8440	-7471B	PPB	11 Feb 2022 14:41:53	U	SMPL
CCB079 - 1	CCB079	-40	-0.0068	-7471B	PPB	11 Feb 2022 14:44:10	U	SMPL
N1457-01DLX10 - 1	TP-1	33805	5.9360	-7471B	PPB	11 Feb 2022 14:46:27	U	SMPL
N1457-03DLX20 - 1	TP-2	106413	18.6851	-7471B	PPB	11 Feb 2022 14:48:44	U	SMPL
N1457-04DLX20 - 1	TP-3	27125	4.7631	-7471B	PPB	11 Feb 2022 14:51:00	U	SMPL
N1457-07DLX10 - 1	TP-4	29311	5.1469	-7471B	PPB	11 Feb 2022 14:53:18	U	SMPL
N1457-09DLX5 - 1	TP-5	26611	4.6728	-7471B	PPB	11 Feb 2022 14:55:33	U	SMPL
N1457-03DL2X50 - 1	TP-2	42659	7.4906	-7471B	PPB	11 Feb 2022 14:57:49	U	SMPL
CCV080 - 1	CCV080	28330	4.9746	-7471B	PPB	11 Feb 2022 15:00:05	U	SMPL
CCB080 - 1	CCB080	-31	-0.0052	-7471B	PPB	11 Feb 2022 15:02:21	U	SMPL