



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118649

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/11/2022	16:40
2	CAL2	1	Water	NA	NA	20.4	7.00	02/11/2022	16:41
3	CAL3	1	Water	NA	NA	20.4	10.03	02/11/2022	16:42
4	ICV	1	Water	NA	NA	20.4	6.99	02/11/2022	16:44
5	CCV1	1	Water	NA	NA	20.4	2.01	02/11/2022	16:45
6	N1517-03	1	Water	NA	NA	20.4	8.27	02/11/2022	16:47
7	N1517-03DUP	1	Water	NA	NA	20.5	8.28	02/11/2022	16:48
8	N1517-04	1	Water	NA	NA	20.2	5.97	02/11/2022	16:49
9	CCV2	1	Water	NA	NA	20.3	12.02	02/11/2022	16:50

118649

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1515 WorkList ID : 156773 Department : Wet-Chemistry Date : 02-11-2022 16:29:34

Due Date		Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location		Customer Sample	Collect Date	Method
02/11/2022		Water	N1517-03	pH	A	Cool 4 deg C	PSEG03	M11	30971	02/11/2022	9040C
02/11/2022		Water	N1517-04	pH		Cool 4 deg C	PSEG03	M11	20991	02/11/2022	9040C

Date/Time 02.11.22 16:30
Raw Sample Received by: JN (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 02.11.22 18:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JN (WC)