



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118650

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/11/2022	16:40
2	CAL2	1	Water	NA	NA	20.4	7.00	02/11/2022	16:41
3	CAL3	1	Water	NA	NA	20.4	10.03	02/11/2022	16:42
4	ICV	1	Water	NA	NA	20.4	6.99	02/11/2022	16:44
5	CCV1	1	Water	NA	NA	20.3	2.02	02/11/2022	17:00
6	N1515-02	1	Solid	20.03	20	20.4	8.73	02/11/2022	17:02
7	N1524-01	1	Solid	20.02	20	20.4	8.66	02/11/2022	17:05
8	N1524-01DUP	1	Solid	20.03	20	20.5	8.67	02/11/2022	17:06
9	CCV2	1	Water	NA	NA	20.5	12.03	02/11/2022	17:09

WORKLIST(Hardcopy Internal Chain)

VB 118150

WorkList Name : CORROSIONITY N1525 WorkList ID : 156774 Department : Wet-Chemistry Date : 02-11-2022 16:30:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/11/2022	Solid	N1515-02	Corrosivity	Cool 4 deg C	PSEG03	M11	OD-1	02/11/2022	9045D
02/11/2022	Solid	N1524-01	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-1	02/11/2022	9045D

Date/Time 02.11.22 16:32
 Raw Sample Received by: JN (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 02.11.22 18:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JN (WC)