



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118651

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/11/2022	16:40
2	CAL2	1	Water	NA	NA	20.4	7.00	02/11/2022	16:41
3	CAL3	1	Water	NA	NA	20.4	10.03	02/11/2022	16:42
4	ICV	1	Water	NA	NA	20.4	6.99	02/11/2022	16:44
5	CCV1	1	Water	NA	NA	20.3	2.02	02/11/2022	17:20
6	N1515-05	1	Water	NA	NA	20.6	9.25	02/11/2022	17:22
7	N1515-05DUP	1	Water	NA	NA	20.7	9.26	02/11/2022	17:23
8	CCV2	1	Water	NA	NA	20.4	12.03	02/11/2022	17:25

WB 11865)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1515 w WorkList ID : 156775 Department : Wet-Chemistry Date : 02-11-2022 16:31:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
02/11/2022	Water	N1515-05	Corrosivity	Cool 4 deg C	PSEG03	M11	OW-1	02/11/2022	9040C

Date/Time 02.11.22 16:32
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: JH (WC)

Date/Time 02.11.22 18:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JH (WC)