



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/14/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:00
In Date: 02/12/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 02/13/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118653

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1503-01	BGZA8	1	1.14	8.45	9.59	8.88	91.6	
N1503-02	BGZA9	2	1.19	8.54	9.73	8.98	91.2	
N1503-03	BGZB0	3	1.15	8.80	9.95	9.42	94.0	
N1503-04	BGZB1	4	1.19	8.46	9.65	8.98	92.1	
N1503-05	BGZB2	5	1.16	8.77	9.93	9.62	96.5	
N1503-06	BGZB3	6	1.19	8.37	9.56	9.00	93.3	
N1503-07	BGZB4	7	1.15	8.81	9.96	9.63	96.3	
N1503-08	BGZB5	8	1.12	8.66	9.78	9.03	91.3	
N1503-09	BGZB6	9	1.15	8.80	9.95	9.64	96.5	
N1503-10	BGZB7	10	1.16	8.82	9.98	9.36	93.0	
N1503-11	BGZB8	11	1.15	8.73	9.88	8.74	86.9	
N1503-12	BGZB9	12	1.19	8.41	9.6	8.72	89.5	
N1503-13	BGZC0	13	1.17	8.51	9.68	8.65	87.9	
N1503-14	BGZC1	14	1.19	8.61	9.8	9.12	92.1	
N1503-15	BGZC2	15	1.14	8.80	9.94	8.98	89.1	
N1503-16	BGZC3	16	1.18	8.64	9.82	8.66	86.6	
N1503-17	BGZC5	17	1.18	8.78	9.96	9.31	92.6	
N1503-18	BGZC5D	18	1.18	8.78	9.96	9.31	92.6	
N1503-19	BGZC5S	19	1.18	8.78	9.96	9.31	92.6	
N1503-20	BGZC6	20	1.19	8.48	9.67	8.92	91.2	
N1503-21	BGZC7	21	1.13	8.70	9.83	9.19	92.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$