

LB118657


INSTRUMENT ID : CV1

SFAM01.1

Linear

μ Abs.:

47329



A= 0.0000e+000

B= 2.1108e-004 *slope*

C= -4.0932e-002 *intercept*

Rho= 0.9999235

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	-0.031	-0.031	48	0.000	48					=
0.05	0.050					205	0				-15
0.20	0.200	0.170	-0.030	997	0.0 %	997					1
2.50	2.500	2.525	0.025	12155	0.0 %	12155					2
5.00	5.000	5.086	0.086	24290	0.0 %	24290					0
7.50	7.500	7.501	0.001	35731	0.0 %	35731					-1
10.0	10.000	9.949	-0.051	47329	0.0 %	47329					

LB118657
INSTRUMENT ID:CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	48	-		0 SFAM01.1	PPB	2/13/2022 13:52	Std
	0.2 S01	997	-		0.2 SFAM01.1	PPB	2/13/2022 13:54	Std
	2.5 S02	12155	-		2.5 SFAM01.1	PPB	2/13/2022 13:57	Std
	5 S03	24290	-		5 SFAM01.1	PPB	2/13/2022 13:59	Std
	7.5 S04	35731	-		7.5 SFAM01.1	PPB	2/13/2022 14:01	Std
	10 S05	47329	-		10 SFAM01.1	PPB	2/13/2022 14:04	Std
ICV90	ICV90	20929	4.3767	-	SFAM01.1	PPB	2/13/2022 14:07	SMPL
ICB90	ICB90	-25	-0.0462	-	SFAM01.1	PPB	2/13/2022 14:09	SMPL
CCV001	CCV001	22417	4.6908	-	SFAM01.1	PPB	2/13/2022 14:12	SMPL
CCB001	CCB001	-25	-0.0462	-	SFAM01.1	PPB	2/13/2022 14:14	SMPL
PB142628BL	PBS628	17	-0.0373	-	SFAM01.1	PPB	2/13/2022 14:16	SMPL
N1408-01	A5700	271	0.0163	-	SFAM01.1	PPB	2/13/2022 14:18	SMPL
N1408-02	A5701	282	0.0186	-	SFAM01.1	PPB	2/13/2022 14:21	SMPL
N1408-03	A5702	181	-0.0027	-	SFAM01.1	PPB	2/13/2022 14:23	SMPL
N1408-04	A5703	245	0.0108	-	SFAM01.1	PPB	2/13/2022 14:25	SMPL
N1408-05	A5704	553	0.0758	-	SFAM01.1	PPB	2/13/2022 14:28	SMPL
N1408-06	A5705	235	0.0087	-	SFAM01.1	PPB	2/13/2022 14:30	SMPL
N1408-07	A5706	455	0.0551	-	SFAM01.1	PPB	2/13/2022 14:32	SMPL
N1408-08	A5707	1592	0.2951	-	SFAM01.1	PPB	2/13/2022 14:34	SMPL
N1408-09	A5708	269	0.0158	-	SFAM01.1	PPB	2/13/2022 14:37	SMPL
N1408-10	A5709	235	0.0087	-	SFAM01.1	PPB	2/13/2022 14:39	SMPL
N1408-14	A5711	785	0.1248	-	SFAM01.1	PPB	2/13/2022 14:48	SMPL
N1408-15	A5712	221	0.0057	-	SFAM01.1	PPB	2/13/2022 14:50	SMPL
N1408-16	A5713	309	0.0243	-	SFAM01.1	PPB	2/13/2022 14:53	SMPL
N1408-17	A5714	206	0.0026	-	SFAM01.1	PPB	2/13/2022 14:55	SMPL
N1408-18	A5715	341	0.031	-	SFAM01.1	PPB	2/13/2022 14:57	SMPL
N1408-19	A5716	381	0.0395	-	SFAM01.1	PPB	2/13/2022 14:59	SMPL
CCV002	CCV002	23908	5.0055	-	SFAM01.1	PPB	2/13/2022 15:02	SMPL
CCB002	CCB002	-37	-0.0487	-	SFAM01.1	PPB	2/13/2022 15:04	SMPL
N1408-20	A5717	462	0.0566	-	SFAM01.1	PPB	2/13/2022 15:06	SMPL
N1408-21	A5718	221	0.0057	-	SFAM01.1	PPB	2/13/2022 15:08	SMPL
N1408-22	A5719	1795	0.338	-	SFAM01.1	PPB	2/13/2022 15:11	SMPL
N1408-23	A5733	60722	12.7762	-	SFAM01.1	PPB	2/13/2022 15:13	SMPL
PB142629BL	PBS629	-73	-0.0563	-	SFAM01.1	PPB	2/13/2022 15:15	SMPL
N1409-01	A5720	131	-0.0133	-	SFAM01.1	PPB	2/13/2022 15:17	SMPL
N1409-02	A5721	190	-0.0008	-	SFAM01.1	PPB	2/13/2022 15:20	SMPL
N1409-03	A5722	1487	0.2729	-	SFAM01.1	PPB	2/13/2022 15:22	SMPL
N1409-04	A5722D	1499	0.2755	-	SFAM01.1	PPB	2/13/2022 15:24	SMPL
N1409-06	A5723	197	0.0007	-	SFAM01.1	PPB	2/13/2022 15:29	SMPL
N1409-07	A5724	1276	0.2284	-	SFAM01.1	PPB	2/13/2022 15:31	SMPL
N1409-08	A5725	13336	2.774	-	SFAM01.1	PPB	2/13/2022 15:33	SMPL
N1409-09	A5726	792	0.1262	-	SFAM01.1	PPB	2/13/2022 15:36	SMPL
N1409-10	A5727	2007	0.3827	-	SFAM01.1	PPB	2/13/2022 15:38	SMPL
N1409-11	A5728	782	0.1241	-	SFAM01.1	PPB	2/13/2022 15:40	SMPL
N1409-12	A5729	855	0.1395	-	SFAM01.1	PPB	2/13/2022 15:42	SMPL
N1409-13	A5732	80538	16.959	-	SFAM01.1	PPB	2/13/2022 15:45	SMPL

LB118657

INSTRUMENT ID:CV1

PB142630BL	PBS630	221	0.0057 -	SFAM01.1	PPB	2/13/2022 15:47	SMPL
N1413-01	YAWK2	2334	0.4517 -	SFAM01.1	PPB	2/13/2022 15:49	SMPL
CCV003	CCV003	24261	5.0801 -	SFAM01.1	PPB	2/13/2022 15:55	SMPL
CCB003	CCB003	-34	-0.0481 -	SFAM01.1	PPB	2/13/2022 15:57	SMPL
N1413-02	YAWK3	3156	0.6252 -	SFAM01.1	PPB	2/13/2022 16:00	SMPL
N1413-03	YAWK4	12907	2.6835 -	SFAM01.1	PPB	2/13/2022 16:02	SMPL
N1413-04	YAWK5	7130	1.4641 -	SFAM01.1	PPB	2/13/2022 16:04	SMPL
N1413-05	YAWK6	2391	0.4638 -	SFAM01.1	PPB	2/13/2022 16:06	SMPL
N1413-06	YAWK7	1573	0.2911 -	SFAM01.1	PPB	2/13/2022 16:09	SMPL
N1413-07	YAWK8	3486	0.6949 -	SFAM01.1	PPB	2/13/2022 16:11	SMPL
N1413-08	YAWK9	2684	0.5256 -	SFAM01.1	PPB	2/13/2022 16:13	SMPL
N1413-09	YAWL0	1717	0.3215 -	SFAM01.1	PPB	2/13/2022 16:15	SMPL
N1413-10	YAWL1	1848	0.3491 -	SFAM01.1	PPB	2/13/2022 16:18	SMPL
N1413-11	YAWL2	1957	0.3721 -	SFAM01.1	PPB	2/13/2022 16:20	SMPL
N1413-12	YAWL3	1774	0.3335 -	SFAM01.1	PPB	2/13/2022 16:22	SMPL
N1413-13	YAWL4	1999	0.381 -	SFAM01.1	PPB	2/13/2022 16:25	SMPL
N1413-14	YAWL5	2180	0.4192 -	SFAM01.1	PPB	2/13/2022 16:27	SMPL
N1413-15	YAWL6	2152	0.4133 -	SFAM01.1	PPB	2/13/2022 16:29	SMPL
N1413-16	YAWL7	2799	0.5499 -	SFAM01.1	PPB	2/13/2022 16:31	SMPL
N1413-17	YAWL8	2579	0.5034 -	SFAM01.1	PPB	2/13/2022 16:34	SMPL
N1413-18	YAWL9	2771	0.544 -	SFAM01.1	PPB	2/13/2022 16:36	SMPL
N1413-19	YAWM0	2070	0.396 -	SFAM01.1	PPB	2/13/2022 16:38	SMPL
N1413-20	YAWM0D	2087	0.3996 -	SFAM01.1	PPB	2/13/2022 16:40	SMPL
N1413-21	YAWM0S	13373	2.7818 -	SFAM01.1	PPB	2/13/2022 16:43	SMPL
CCV004	CCV004	25177	5.2734 -	SFAM01.1	PPB	2/13/2022 16:45	SMPL
CCB004	CCB004	-22	-0.0456 -	SFAM01.1	PPB	2/13/2022 16:47	SMPL
N1413-22	YAWM1	2561	0.4996 -	SFAM01.1	PPB	2/13/2022 16:50	SMPL
PB142631BL	PBS631	144	-0.0105 -	SFAM01.1	PPB	2/13/2022 16:52	SMPL
N1414-01	YAWM2	11543	2.3956 -	SFAM01.1	PPB	2/13/2022 16:54	SMPL
N1414-02	YAWM3	2239	0.4317 -	SFAM01.1	PPB	2/13/2022 16:56	SMPL
N1414-03	YAWM4	2399	0.4654 -	SFAM01.1	PPB	2/13/2022 16:59	SMPL
N1414-04	YAWM5	3538	0.7059 -	SFAM01.1	PPB	2/13/2022 17:01	SMPL
N1414-05	YAWZ4	3685	0.7369 -	SFAM01.1	PPB	2/13/2022 17:03	SMPL
N1414-06	YAWZ5	2923	0.5761 -	SFAM01.1	PPB	2/13/2022 17:05	SMPL
N1414-07	YAWZ6	33856	7.1054 -	SFAM01.1	PPB	2/13/2022 17:08	SMPL
N1414-08	YAWZ7	3175	0.6292 -	SFAM01.1	PPB	2/13/2022 17:10	SMPL
N1414-09	YAWZ8	3223	0.6394 -	SFAM01.1	PPB	2/13/2022 17:12	SMPL
N1414-10	YAWZ9	3473	0.6921 -	SFAM01.1	PPB	2/13/2022 17:15	SMPL
N1414-11	YAX00	2140	0.4108 -	SFAM01.1	PPB	2/13/2022 17:17	SMPL
N1414-12	YAX01	2087	0.3996 -	SFAM01.1	PPB	2/13/2022 17:19	SMPL
N1414-13	YAX24	3477	0.693 -	SFAM01.1	PPB	2/13/2022 17:21	SMPL
N1414-14	YAX25	3351	0.6664 -	SFAM01.1	PPB	2/13/2022 17:24	SMPL
N1414-15	YAX26	2534	0.4939 -	SFAM01.1	PPB	2/13/2022 17:26	SMPL
N1414-16	YAX27	2962	0.5843 -	SFAM01.1	PPB	2/13/2022 17:28	SMPL
N1414-17	YAX28	2923	0.5761 -	SFAM01.1	PPB	2/13/2022 17:31	SMPL
N1414-18	YAX29	6668	1.3665 -	SFAM01.1	PPB	2/13/2022 17:33	SMPL
CCV005	CCV005	23582	4.9367 -	SFAM01.1	PPB	2/13/2022 17:35	SMPL

LB118657

INSTRUMENT ID:CV1

CCB005	CCB005	-18	-0.0447 -	SFAM01.1	PPB	2/13/2022 17:37 SMPL
N1414-19	YAX30	3523	0.7027 -	SFAM01.1	PPB	2/13/2022 17:40 SMPL
N1414-20	YAX31	3364	0.6691 -	SFAM01.1	PPB	2/13/2022 17:42 SMPL
N1414-21	YAX31D	3366	0.6696 -	SFAM01.1	PPB	2/13/2022 17:44 SMPL
N1414-22	YAX31S	12477	2.5927 -	SFAM01.1	PPB	2/13/2022 17:47 SMPL
PB142632BL	PBS632	49	-0.0306 -	SFAM01.1	PPB	2/13/2022 17:49 SMPL
N1437-01	YAX48	3230	0.6409 -	SFAM01.1	PPB	2/13/2022 17:51 SMPL
N1437-02	YAX49	35775	7.5104 -	SFAM01.1	PPB	2/13/2022 17:53 SMPL
N1437-03	YAX50	26311	5.5128 -	SFAM01.1	PPB	2/13/2022 17:56 SMPL
N1437-04	YAX51	30719	6.4432 -	SFAM01.1	PPB	2/13/2022 17:58 SMPL
N1437-05	YAX52	8139	1.677 -	SFAM01.1	PPB	2/13/2022 18:00 SMPL
N1437-06	YAX53	4610	0.9321 -	SFAM01.1	PPB	2/13/2022 18:02 SMPL
N1437-07	YAX54	4040	0.8118 -	SFAM01.1	PPB	2/13/2022 18:05 SMPL
N1437-08	YAX55	4033	0.8104 -	SFAM01.1	PPB	2/13/2022 18:07 SMPL
N1437-09	YAWN4	2460	0.4783 -	SFAM01.1	PPB	2/13/2022 18:09 SMPL
N1437-10	YAWN5	16375	3.4155 -	SFAM01.1	PPB	2/13/2022 18:12 SMPL
N1437-11	YAWN6	3964	0.7958 -	SFAM01.1	PPB	2/13/2022 18:14 SMPL
N1437-12	YAWN7	4675	0.9459 -	SFAM01.1	PPB	2/13/2022 18:16 SMPL
N1437-13	YAWN8	6371	1.3039 -	SFAM01.1	PPB	2/13/2022 18:18 SMPL
N1437-14	YAWN9	9419	1.9472 -	SFAM01.1	PPB	2/13/2022 18:21 SMPL
N1437-15	YAWN9D	9292	1.9204 -	SFAM01.1	PPB	2/13/2022 18:23 SMPL
CCV006	CCV006	24207	5.0687 -	SFAM01.1	PPB	2/13/2022 18:25 SMPL
CCB006	CCB006	-14	-0.0439 -	SFAM01.1	PPB	2/13/2022 18:28 SMPL
N1437-16	YAWN9S	18159	3.7921 -	SFAM01.1	PPB	2/13/2022 18:30 SMPL
N1437-17	YAWP0	2793	0.5486 -	SFAM01.1	PPB	2/13/2022 18:32 SMPL
N1437-18	YAWP1	3333	0.6626 -	SFAM01.1	PPB	2/13/2022 18:34 SMPL
N1437-19	YAWM6	11305	2.3453 -	SFAM01.1	PPB	2/13/2022 18:37 SMPL
N1437-20	YAWM7	5980	1.2213 -	SFAM01.1	PPB	2/13/2022 18:39 SMPL
N1437-21	YAWM8	2971	0.5862 -	SFAM01.1	PPB	2/13/2022 18:41 SMPL
N1437-22	YAWM9	4138	0.8325 -	SFAM01.1	PPB	2/13/2022 18:44 SMPL
PB142633BL	PBS633	247	0.0112 -	SFAM01.1	PPB	2/13/2022 18:46 SMPL
N1450-01	YAWN0	2541	0.4954 -	SFAM01.1	PPB	2/13/2022 18:48 SMPL
N1450-02	YAWN1	2989	0.59 -	SFAM01.1	PPB	2/13/2022 18:50 SMPL
N1450-03	YAWN2	3009	0.5942 -	SFAM01.1	PPB	2/13/2022 18:53 SMPL
N1450-04	YAWN3	2669	0.5224 -	SFAM01.1	PPB	2/13/2022 18:55 SMPL
N1450-05	YAWP2	2662	0.521 -	SFAM01.1	PPB	2/13/2022 18:57 SMPL
N1450-06	YAWP3	3638	0.727 -	SFAM01.1	PPB	2/13/2022 19:00 SMPL
N1450-07	YAWP4	3766	0.754 -	SFAM01.1	PPB	2/13/2022 19:02 SMPL
N1450-08	YAWP5	4030	0.8097 -	SFAM01.1	PPB	2/13/2022 19:04 SMPL
N1450-09	YAWP6	3740	0.7485 -	SFAM01.1	PPB	2/13/2022 19:06 SMPL
N1450-10	YAWP7	16030	3.3427 -	SFAM01.1	PPB	2/13/2022 19:09 SMPL
N1450-11	YAWQ5	775	0.1227 -	SFAM01.1	PPB	2/13/2022 19:11 SMPL
N1450-12	YAWQ6	1052	0.1811 -	SFAM01.1	PPB	2/13/2022 19:13 SMPL
CCV007	CCV007	25828	5.4108 -	SFAM01.1	PPB	2/13/2022 19:16 SMPL
CCB007	CCB007	-18	-0.0447 -	SFAM01.1	PPB	2/13/2022 19:18 SMPL
N1450-13	YAWQ7	1149	0.2016 -	SFAM01.1	PPB	2/13/2022 19:20 SMPL
N1450-14	YAWQ8	983	0.1666 -	SFAM01.1	PPB	2/13/2022 19:22 SMPL

LB118657

INSTRUMENT ID:CV1

N1450-15	YAWQ8D	971	0.164 -	SFAM01.1	PPB	2/13/2022 19:25 SMPL
N1450-16	YAWQ8S	12460	2.5891 -	SFAM01.1	PPB	2/13/2022 19:27 SMPL
N1450-17	YAWQ9	1595	0.2957 -	SFAM01.1	PPB	2/13/2022 19:29 SMPL
N1450-18	YAWR0	1702	0.3183 -	SFAM01.1	PPB	2/13/2022 19:32 SMPL
N1450-19	YAWR1	1951	0.3709 -	SFAM01.1	PPB	2/13/2022 19:34 SMPL
N1450-20	YAWR2	1758	0.3301 -	SFAM01.1	PPB	2/13/2022 19:36 SMPL
N1450-21	YAWR3	1826	0.3445 -	SFAM01.1	PPB	2/13/2022 19:38 SMPL
N1450-22	YAWR4	1686	0.3149 -	SFAM01.1	PPB	2/13/2022 19:41 SMPL
PB142634BL	PBS634	184	-0.0021 -	SFAM01.1	PPB	2/13/2022 19:43 SMPL
N1451-01	YAWP8	2989	0.59 -	SFAM01.1	PPB	2/13/2022 19:45 SMPL
N1451-02	YAWP9	3748	0.7502 -	SFAM01.1	PPB	2/13/2022 19:48 SMPL
N1451-03	YAWQ0	2457	0.4777 -	SFAM01.1	PPB	2/13/2022 19:50 SMPL
N1451-04	YAWQ1	4431	0.8944 -	SFAM01.1	PPB	2/13/2022 19:52 SMPL
N1451-05	YAWQ2	2391	0.4638 -	SFAM01.1	PPB	2/13/2022 19:55 SMPL
N1451-06	YAWQ3	2649	0.5182 -	SFAM01.1	PPB	2/13/2022 19:57 SMPL
N1451-07	YAWQ4	3329	0.6618 -	SFAM01.1	PPB	2/13/2022 19:59 SMPL
N1451-08	YAX23	4708	0.9528 -	SFAM01.1	PPB	2/13/2022 20:01 SMPL
N1451-09	YAX23D	4639	0.9383 -	SFAM01.1	PPB	2/13/2022 20:04 SMPL
CCV008	CCV008	25242	5.2871 -	SFAM01.1	PPB	2/13/2022 20:06 SMPL
CCB008	CCB008	-8	-0.0426 -	SFAM01.1	PPB	2/13/2022 20:08 SMPL
N1451-10	YAX23S	16194	3.3773 -	SFAM01.1	PPB	2/13/2022 20:11 SMPL
N1451-11	YAWR5	1078	0.1866 -	SFAM01.1	PPB	2/13/2022 20:13 SMPL
N1451-12	YAWR6	3603	0.7196 -	SFAM01.1	PPB	2/13/2022 20:15 SMPL
N1451-13	YAWR7	2216	0.4268 -	SFAM01.1	PPB	2/13/2022 20:17 SMPL
N1451-14	YAWR8	1193	0.2109 -	SFAM01.1	PPB	2/13/2022 20:20 SMPL
N1451-15	YAWR9	1239	0.2206 -	SFAM01.1	PPB	2/13/2022 20:22 SMPL
N1451-16	YAWS0	1111	0.1936 -	SFAM01.1	PPB	2/13/2022 20:24 SMPL
N1451-17	YAWS1	1164	0.2048 -	SFAM01.1	PPB	2/13/2022 20:27 SMPL
N1451-18	YAWS2	1433	0.2615 -	SFAM01.1	PPB	2/13/2022 20:29 SMPL
N1451-19	YAWS3	2937	0.579 -	SFAM01.1	PPB	2/13/2022 20:31 SMPL
N1451-20	YAWS4	1960	0.3728 -	SFAM01.1	PPB	2/13/2022 20:34 SMPL
N1451-21	YAWS5	1568	0.29 -	SFAM01.1	PPB	2/13/2022 20:36 SMPL
N1451-22	YAWS6	4073	0.8188 -	SFAM01.1	PPB	2/13/2022 20:38 SMPL
N1408-11	A5710	686	0.1039 -	SFAM01.1	PPB	2/13/2022 20:40 SMPL
N1408-12	A5710D	699	0.1066 -	SFAM01.1	PPB	2/13/2022 20:43 SMPL
N1408-13	A5710S	13180	2.7411 -	SFAM01.1	PPB	2/13/2022 20:45 SMPL
N1409-05	A5722S	14085	2.9321 -	SFAM01.1	PPB	2/13/2022 20:47 SMPL
N1408-23DLX2	A5733	29706	6.2294 -	SFAM01.1	PPB	2/13/2022 20:50 SMPL
N1409-13DLX2	A5732	35999	7.5577 -	SFAM01.1	PPB	2/13/2022 20:52 SMPL
CCV009	CCV009	24782	5.19 -	SFAM01.1	PPB	2/13/2022 20:54 SMPL
CCB009	CCB009	-16	-0.0443 -	SFAM01.1	PPB	2/13/2022 20:56 SMPL