



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118661

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/14/2022	10:26
2	CAL2	1	Water	NA	NA	20.3	7.00	02/14/2022	10:27
3	CAL3	1	Water	NA	NA	20.3	10.03	02/14/2022	10:29
4	ICV	1	Water	NA	NA	20.6	6.99	02/14/2022	10:30
5	CCV1	1	Water	NA	NA	20.4	2.01	02/14/2022	10:31
6	N1529-01	1	Water	NA	NA	20.4	5.61	02/14/2022	10:35
7	N1529-01DUP	1	Water	NA	NA	20.5	5.62	02/14/2022	10:36
8	N1529-02	1	Water	NA	NA	20.6	5.15	02/14/2022	10:37
9	CCV2	1	Water	NA	NA	20.4	12.02	02/14/2022	10:40

118661

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1529 WorkList ID : 15818 Department : Wet-Chemistry Date : 02-14-2022 10:23:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/08/2022	Water	N1529-01	pH <u>A</u>	Cool 4 deg C	STER14	N21	JS10000	02/08/2022	9040C
02/08/2022	Water	N1529-02	pH <u>A</u>	Cool 4 deg C	STER14	N21	MINI-CELL	02/08/2022	9040C

Date/Time 02-14-22 10:25
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)

Date/Time 02-14-22 11:30
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: JP (WC)