

LB118664


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

5022E



A= 0.0000e+000

B= 2.0014e-004

C= -3.8435e-002

Rho= 0.9997025

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.038	-0.038	0	0.000	0					-
0.2	0.200	0.174	-0.026	1061	0.0 %	1061					-13
2.5	2.500	2.533	0.033	12847	0.0 %	12847					1
5.0	5.000	5.159	0.159	25970	0.0 %	25970					3
7.5	7.500	7.359	-0.141	36962	0.0 %	36962					-2
10.0	10.000	10.014	0.014	50226	0.0 %	50226					0

LB118664 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Feb 2022 13:06:18

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	0	-	0.0000	7471B	PPB	14 Feb 2022 13:40:15	S	Std
0.2 - 1	50.2	1061	-	0.2000	7471B	PPB	14 Feb 2022 13:42:33	S	Std
2.5 - 1	52.5	12847	-	2.5000	7471B	PPB	14 Feb 2022 13:44:50	S	Std
5.0 - 1	55	25970	-	5.0000	7471B	PPB	14 Feb 2022 13:47:07	S	Std
7.5 - 1	57.5	36962	-	7.5000	7471B	PPB	14 Feb 2022 13:49:25	S	Std
10.0 - 1	510	50226	-	10.0000	7471B	PPB	14 Feb 2022 13:54:34	S	Std
ICV87 - 1	ICV87	19845	3.9333	-	7471B	PPB	14 Feb 2022 13:57:13	U	SMPL
ICB87 - 1	ICB87	-50	-0.0484	-	7471B	PPB	14 Feb 2022 13:59:28	U	SMPL
CCV276 - 1	CCV276	24057	4.7763	-	7471B	PPB	14 Feb 2022 14:01:44	U	SMPL
CCB276 - 1	CCB276	-56	-0.0496	-	7471B	PPB	14 Feb 2022 14:03:59	U	SMPL
CRA - 1	CRA	1070	0.1757	-	7471B	PPB	14 Feb 2022 14:06:15	U	SMPL
HighStd - 1	HighStd	51312	10.2310	-	7471B	PPB	14 Feb 2022 14:08:30	U	SMPL
ChkStd - 1	ChkStd	35812	7.1289	-	7471B	PPB	14 Feb 2022 14:10:46	U	SMPL
PB142626BL - 1	PBS	-6	-0.0396	-	7471B	PPB	14 Feb 2022 14:15:37	U	SMPL
PB142626BS - 1	LCSS	18646	3.6933	-	7471B	PPB	14 Feb 2022 14:17:53	U	SMPL
N1477-01 - 1	20220079-BERKELEY-7-COMP	4880	0.9382	-	7471B	PPB	14 Feb 2022 14:20:09	U	SMPL
N1477-02 - 1	20220080-BERKELEY-8-COMP	7576	1.4778	-	7471B	PPB	14 Feb 2022 14:22:26	U	SMPL
N1477-03 - 1	20220087-AMENDED-BERKELEY-7-COMP	5293	1.0209	-	7471B	PPB	14 Feb 2022 14:24:43	U	SMPL
N1477-04 - 1	20220088-AMENDED-BERKELEY-8-COMP	6869	1.3363	-	7471B	PPB	14 Feb 2022 14:27:00	U	SMPL
N1487-01 - 1	251488	21782	4.3209	-	7471B	PPB	14 Feb 2022 14:29:15	U	SMPL
CCV277 - 1	CCV277	23645	4.6938	-	7471B	PPB	14 Feb 2022 14:31:32	U	SMPL
CCB277 - 1	CCB277	-34	-0.0452	-	7471B	PPB	14 Feb 2022 14:33:48	U	SMPL
N1489-01 - 1	COMP-1	5935	1.1494	-	7471B	PPB	14 Feb 2022 14:36:03	U	SMPL
N1489-02 - 1	COMP-2	6993	1.3611	-	7471B	PPB	14 Feb 2022 14:38:19	U	SMPL
N1489-03 - 1	COMP-3	6594	1.2813	-	7471B	PPB	14 Feb 2022 14:40:35	U	SMPL
N1494-01 - 1	ASTORIA-YARD-PIT-10	68349	13.6407	-	7471B	PPB	14 Feb 2022 14:42:51	U	SMPL
N1494-02 - 1	21ST-STREET-LAUNCH-PIT-3	1565	0.2748	-	7471B	PPB	14 Feb 2022 14:45:07	U	SMPL
N1495-01 - 1	NB-08-021022	1079	0.1775	-	7471B	PPB	14 Feb 2022 14:47:25	U	SMPL
N1495-01DUP - 1	NB-08-021022DUP	1077	0.1771	-	7471B	PPB	14 Feb 2022 14:52:18	U	SMPL
N1495-01MS - 1	NB-08-021022MS	21439	4.2523	-	7471B	PPB	14 Feb 2022 14:54:35	U	SMPL
N1495-01MSD - 1	NB-08-021022MSD	21522	4.2689	-	7471B	PPB	14 Feb 2022 14:56:52	U	SMPL
N1496-01 - 1	OK-01-021022	1391	0.2400	-	7471B	PPB	14 Feb 2022 14:59:08	U	SMPL
CCV278 - 1	CCV278	24118	4.7885	-	7471B	PPB	14 Feb 2022 15:01:24	U	SMPL
CCB278 - 1	CCB278	-46	-0.0476	-	7471B	PPB	14 Feb 2022 15:03:40	U	SMPL
N1505-01 - 1	SU-02-021022	3521	0.6662	-	7471B	PPB	14 Feb 2022 15:05:56	U	SMPL
N1505-03 - 1	SU-03-021022	1256	0.2129	-	7471B	PPB	14 Feb 2022 15:08:11	U	SMPL
N1511-01 - 1	20220034-NACOTE-1-COMP	1788	0.3194	-	7471B	PPB	14 Feb 2022 15:10:27	U	SMPL
N1511-02 - 1	20220041-AMENDED-NACOTE-1-COMP	1393	0.2404	-	7471B	PPB	14 Feb 2022 15:12:43	U	SMPL
N1515-01 - 1	OD-1	9025	1.7678	-	7471B	PPB	14 Feb 2022 15:14:58	U	SMPL
N1518-01 - 1	OR-02-021122	2156	0.3931	-	7471B	PPB	14 Feb 2022 15:17:15	U	SMPL
N1524-01 - 1	TP-1	20704	4.1052	-	7471B	PPB	14 Feb 2022 15:19:32	U	SMPL
N1525-01 - 1	OR-02-021122	1585	0.2788	-	7471B	PPB	14 Feb 2022 15:21:49	U	SMPL
PB142627BL - 1	PBS	5	-0.0374	-	7471B	PPB	14 Feb 2022 15:24:06	U	SMPL
PB142627BS - 1	LCSS	20755	4.1154	-	7471B	PPB	14 Feb 2022 15:26:22	U	SMPL
CCV279 - 1	CCV279	24217	4.8083	-	7471B	PPB	14 Feb 2022 15:28:39	U	SMPL
CCB279 - 1	CCB279	-38	-0.0460	-	7471B	PPB	14 Feb 2022 15:30:55	U	SMPL
N1521-02 - 1	TP1-A	776	0.1169	-	7471B	PPB	14 Feb 2022 15:33:11	U	SMPL
N1521-02DUP - 1	TP1-ADUP	949	0.1515	-	7471B	PPB	14 Feb 2022 15:35:26	U	SMPL
N1521-02MS - 1	TP1-AMS	22000	4.3646	-	7471B	PPB	14 Feb 2022 15:37:42	U	SMPL
N1521-02MSD - 1	TP1-AMSD	20863	4.1370	-	7471B	PPB	14 Feb 2022 15:39:58	U	SMPL
N1521-04 - 1	TP1-B	591	0.0798	-	7471B	PPB	14 Feb 2022 15:42:14	U	SMPL
N1521-06 - 1	TP1-C	23	-0.0338	-	7471B	PPB	14 Feb 2022 15:44:31	U	SMPL
N1521-08 - 1	TP2-A	590	0.0796	-	7471B	PPB	14 Feb 2022 15:46:47	U	SMPL
N1521-10 - 1	TP2-B	89	-0.0206	-	7471B	PPB	14 Feb 2022 15:49:05	U	SMPL
N1521-12 - 1	TP2-C	74	-0.0236	-	7471B	PPB	14 Feb 2022 15:51:22	U	SMPL
N1521-14 - 1	TP3-A	944	0.1505	-	7471B	PPB	14 Feb 2022 15:53:37	U	SMPL
CCV280 - 1	CCV280	26366	5.2384	-	7471B	PPB	14 Feb 2022 15:55:53	U	SMPL
CCB280 - 1	CCB280	-31	-0.0446	-	7471B	PPB	14 Feb 2022 15:58:09	U	SMPL
N1521-16 - 1	TP3-B	570	0.0756	-	7471B	PPB	14 Feb 2022 16:00:25	U	SMPL
N1521-18 - 1	TP3-C	2477	0.4573	-	7471B	PPB	14 Feb 2022 16:02:41	U	SMPL
N1526-01 - 1	HR-1-021122	5469	1.0561	-	7471B	PPB	14 Feb 2022 16:04:57	U	SMPL
N1526-03 - 1	HE-2-021122	1582	0.2782	-	7471B	PPB	14 Feb 2022 16:07:13	U	SMPL
N1495-01LX5 - 1		218	0.0052	-	7471B	PPB	14 Feb 2022 16:09:29	U	SMPL
N1495-01A - 1		21572	4.2789	-	7471B	PPB	14 Feb 2022 16:11:45	U	SMPL
N1521-02LX5 - 1		125	-0.0134	-	7471B	PPB	14 Feb 2022 16:14:02	U	SMPL
N1521-02A - 1		20867	4.1378	-	7471B	PPB	14 Feb 2022 16:16:19	U	SMPL
N1494-01DLX5 - 1		14435	2.8505	-	7471B	PPB	14 Feb 2022 16:18:36	U	SMPL
CCV281 - 1	CCV281	24965	4.9580	-	7471B	PPB	14 Feb 2022 16:20:53	U	SMPL
CCB281 - 1	CCB281	-38	-0.0460	-	7471B	PPB	14 Feb 2022 16:23:10	U	SMPL