

LB118675


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

42585



A= 0.0000e+000

B= 2.3363e-004

C= -3.5883e-002

Rho= 0.9998839

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.020	-0.020	70	0.000	70					-8
0.2	0.200	0.184	-0.016	942	0.0 %	942					-1
2.5	2.500	2.479	-0.021	10763	0.0 %	10763					1
5.0	5.000	5.068	0.068	21847	0.0 %	21847					1
7.5	7.500	7.575	0.075	32579	0.0 %	32579					1
10.0	10.000	9.913	-0.087	42585	0.0 %	42585					-1

LB118675 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 15 Feb 2022 09:31:02							
Sample ID	Extended ID	μ Abs	Conc.	Strnd ConcMethod	Units	Date	Type	Type			
0.0 - 1	50	70	-	0.00007470A	PPB	15 Feb 2022 09:31:24	S	Std			
0.2 - 1	50.2	942	-	0.20007470A	PPB	15 Feb 2022 09:33:41	S	Std			
2.5 - 1	52.5	10763	-	2.50007470A	PPB	15 Feb 2022 09:35:58	S	Std			
5.0 - 1	55	21847	-	5.00007470A	PPB	15 Feb 2022 09:38:16	S	Std			
7.5 - 1	55.5	32579	-	7.50007470A	PPB	15 Feb 2022 09:40:32	S	Std			
10.0 - 1	510	42585	-	10.00007470A	PPB	15 Feb 2022 09:42:48	S	Std			
ICV88 - 1	ICV88	17385	4.0257	-7470A	PPB	15 Feb 2022 09:45:41	U	SMPL			
ICB88 - 1	ICB88	-23	-0.0413	-7470A	PPB	15 Feb 2022 09:47:56	U	SMPL			
CCV82 - 1	CCV82	19877	4.6079	-7470A	PPB	15 Feb 2022 09:52:58	U	SMPL			
CCB82 - 1	CCB82	-31	-0.0431	-7470A	PPB	15 Feb 2022 09:55:13	U	SMPL			
CRA - 1	CRA	839	0.1601	-7470A	PPB	15 Feb 2022 09:57:28	U	SMPL			
HighStd - 1	HighStd	41767	9.7220	-7470A	PPB	15 Feb 2022 09:59:46	U	SMPL			
ChkStd - 1	ChkStd	28977	6.7339	-7470A	PPB	15 Feb 2022 10:02:03	U	SMPL			
PB142673BL - 1	PBW	-55	-0.0487	-7470A	PPB	15 Feb 2022 10:04:21	U	SMPL			
PB142673BS - 1	LCSW	14133	3.2660	-7470A	PPB	15 Feb 2022 10:06:38	U	SMPL			
N1449-03 - 1	SPLP-LEACHATE-AMENDED-COMP	-31	-0.0431	-7470A	PPB	15 Feb 2022 10:08:55	U	SMPL			
N1449-03DUP - 1	SPLP-LEACHATE-AMENDED-COMPDUP	13	-0.0328	-7470A	PPB	15 Feb 2022 10:11:11	U	SMPL			
N1449-03MS - 1	SPLP-LEACHATE-AMENDED-COMPMS	14101	3.2585	-7470A	PPB	15 Feb 2022 10:13:29	U	SMPL			
N1449-03MSD - 1	SPLP-LEACHATE-AMENDED-COMPMSD	15909	3.6809	-7470A	PPB	15 Feb 2022 10:15:46	U	SMPL			
N1477-05 - 1	SPLP-LEACHATE-AMENDED-COMP-7	-31	-0.0431	-7470A	PPB	15 Feb 2022 10:18:02	U	SMPL			
CCV83 - 1	CCV83	20144	4.6703	-7470A	PPB	15 Feb 2022 10:26:03	U	SMPL			
CCB83 - 1	CCB83	-43	-0.0459	-7470A	PPB	15 Feb 2022 10:28:18	U	SMPL			
N1477-06 - 1	SPLP-LEACHATE-AMENDED-COMP-8	12	-0.0331	-7470A	PPB	15 Feb 2022 10:30:33	U	SMPL			
PB142674BL - 1	PB142674BL	15	-0.0324	-7470A	PPB	15 Feb 2022 10:32:49	U	SMPL			
PB142674BS - 1	PB142674BS	16614	3.8456	-7470A	PPB	15 Feb 2022 10:35:05	U	SMPL			
N1470-01 - 1	CTY2-WC-ROLLOFF	-23	-0.0413	-7470A	PPB	15 Feb 2022 10:37:21	U	SMPL			
N1491-01 - 1	SOIL-DRUM	8	-0.0340	-7470A	PPB	15 Feb 2022 10:39:37	U	SMPL			
N1491-01DUP - 1	SOIL-DRUMDUP	24	-0.0303	-7470A	PPB	15 Feb 2022 10:41:54	U	SMPL			
N1491-01MS - 1	SOIL-DRUMMS	16419	3.8000	-7470A	PPB	15 Feb 2022 10:44:11	U	SMPL			
N1491-01MSD - 1	SOIL-DRUMMSD	16410	3.7979	-7470A	PPB	15 Feb 2022 10:46:28	U	SMPL			
PB142675BL - 1	PBW	-28	-0.0424	-7470A	PPB	15 Feb 2022 10:48:45	U	SMPL			
PB142675BS - 1	LCSW	16119	3.7299	-7470A	PPB	15 Feb 2022 10:51:01	U	SMPL			
CCV84 - 1	CCV84	21165	4.9088	-7470A	PPB	15 Feb 2022 10:53:17	U	SMPL			
CCB84 - 1	CCB84	-29	-0.0427	-7470A	PPB	15 Feb 2022 10:55:33	U	SMPL			
N1491-02 - 1	WATER-DRUM-1-3	7	-0.0342	-7470A	PPB	15 Feb 2022 10:57:49	U	SMPL			
N1491-03 - 1	TOTE-17127	20	-0.0312	-7470A	PPB	15 Feb 2022 11:00:05	U	SMPL			
N1491-03DUP - 1	TOTE-17127DUP	22	-0.0307	-7470A	PPB	15 Feb 2022 11:02:21	U	SMPL			
N1491-03MS - 1	TOTE-17127MS	18003	4.1701	-7470A	PPB	15 Feb 2022 11:04:37	U	SMPL			
N1491-03MSD - 1	TOTE-17127MSD	17377	4.0238	-7470A	PPB	15 Feb 2022 11:06:53	U	SMPL			
PB142673TB - 1	PB142591TB	-31	-0.0431	-7470A	PPB	15 Feb 2022 11:09:11	U	SMPL			
PB142589TB - 1	PB142589TB	0	-0.0359	-7470A	PPB	15 Feb 2022 11:11:27	U	SMPL			
PB142600TB - 1	PB142600TB	14	-0.0326	-7470A	PPB	15 Feb 2022 11:13:44	U	SMPL			
CCV85 - 1	CCV85	19772	4.5834	-7470A	PPB	15 Feb 2022 11:16:02	U	SMPL			
CCB85 - 1	CCB85	-31	-0.0431	-7470A	PPB	15 Feb 2022 11:18:18	U	SMPL			
N1449-03LX5 - 1		70	-0.0195	-7470A	PPB	15 Feb 2022 11:20:33	U	SMPL			
N1449-03A - 1		17047	3.9467	-7470A	PPB	15 Feb 2022 11:22:49	U	SMPL			
N1491-01LX5 - 1		61	-0.0216	-7470A	PPB	15 Feb 2022 11:25:05	U	SMPL			
N1491-01A - 1		17040	3.9451	-7470A	PPB	15 Feb 2022 11:27:21	U	SMPL			
N1491-03LX5 - 1		-37	-0.0445	-7470A	PPB	15 Feb 2022 11:29:37	U	SMPL			
N1491-03A - 1		17826	4.1287	-7470A	PPB	15 Feb 2022 11:31:53	U	SMPL			
CCV86 - 1	CCV86	20032	4.6441	-7470A	PPB	15 Feb 2022 11:34:10	U	SMPL			
CCB86 - 1	CCB86	-30	-0.0429	-7470A	PPB	15 Feb 2022 11:36:26	U	SMPL			