



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118676

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	02/15/2022	11:25
2	N1515-05		1	15	100.00	>212.0	>212.0	02/15/2022	11:40
3	N1515-05DUP		1	15	100.00	>212.0	>212.0	02/15/2022	12:05
4	N1517-01		1	14	100.00	>212.0	>212.0	02/15/2022	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118676

WorkList Name : fp21522 WorkList ID : 156869 Department : Wet-Chemistry Date : 02-15-2022 11:01:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/17/2022	Water	N1515-05	Flash Point	Cool 4 deg C	PSEG03	M11	OW-1	02/11/2022	1010B
02/16/2022	Water	N1517-01	Flash Point	Cool 4 deg C	PSEG03	M11	30888	02/11/2022	1010B

Date/Time 02/15/22 11:05
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: R(wc)

Date/Time 02/15/22 12:40
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: KS(wc)