

7470A


LB118678

INSTRUMENT ID: CVL

Linear

 μ Abs.:

44619



A= 0.0000e+000

B= 2.2482e-004

C= -2.1244e-002

Rho= 0.9999774

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.010	-0.010	48	0.000	48					-5
0.2	0.200	0.191	-0.009	942	0.0 %	942					1
2.5	2.500	2.517	0.017	11292	0.0 %	11292					1
5.0	5.000	5.036	0.036	22493	0.0 %	22493					1
7.5	7.500	7.457	-0.043	33265	0.0 %	33265					1
10.0	10.000	10.010	0.010	44619	0.0 %	44619					0

LB118678 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Feb 2022 14:19:02

Sample ID	Extended ID	µ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	48	-	0.00007470A	PPB	15 Feb 2022 14:32:55	S	Std
0.2 - 1	50.2	942	-	0.20007470A	PPB	15 Feb 2022 14:35:12	S	Std
2.5 - 1	50.5	11292	-	2.50007470A	PPB	15 Feb 2022 14:37:29	S	Std
5.0 - 1	50.5	22493	-	5.00007470A	PPB	15 Feb 2022 14:39:47	S	Std
7.5 - 1	50.5	33265	-	7.50007470A	PPB	15 Feb 2022 14:42:03	S	Std
10.0 - 1	50.5	44619	-	10.00007470A	PPB	15 Feb 2022 14:47:22	S	Std
ICV89 - 1	ICV89	17272	3.8618	-7470A	PPB	15 Feb 2022 14:50:11	U	SMPL
ICB89 - 1	ICB89	23	-0.0161	-7470A	PPB	15 Feb 2022 14:52:26	U	SMPL
CCV87 - 1	CCV87	21788	4.8770	-7470A	PPB	15 Feb 2022 14:54:42	U	SMPL
CCB87 - 1	CCB87	20	-0.0167	-7470A	PPB	15 Feb 2022 14:56:58	U	SMPL
CRA - 1	CRA	877	0.1759	-7470A	PPB	15 Feb 2022 14:59:14	U	SMPL
HighStd - 1	HighStd	44744	10.0379	-7470A	PPB	15 Feb 2022 15:01:29	U	SMPL
ChkStd - 1	ChkStd	29719	6.6600	-7470A	PPB	15 Feb 2022 15:06:25	U	SMPL
PB142691BL - 1	PBW	180	0.0192	-7470A	PPB	15 Feb 2022 15:08:41	U	SMPL
PB142691BS - 1	LCSW	16009	3.5778	-7470A	PPB	15 Feb 2022 15:10:58	U	SMPL
N1455-12 - 1	TP-1	243	0.0334	-7470A	PPB	15 Feb 2022 15:13:15	U	SMPL
N1515-02 - 1	OD-1	239	0.0325	-7470A	PPB	15 Feb 2022 15:15:32	U	SMPL
N1532-02 - 1	VNJ-209	221	0.0284	-7470A	PPB	15 Feb 2022 15:17:51	U	SMPL
N1535-02 - 1	368	248	0.0345	-7470A	PPB	15 Feb 2022 15:20:08	U	SMPL
N1535-02DUP - 1	368DUP	231	0.0307	-7470A	PPB	15 Feb 2022 15:22:24	U	SMPL
CCV88 - 1	CCV88	21493	4.8107	-7470A	PPB	15 Feb 2022 15:24:39	U	SMPL
CCB88 - 1	CCB88	25	-0.0156	-7470A	PPB	15 Feb 2022 15:26:54	U	SMPL
N1535-02MS - 1	368MS	16678	3.7282	-7470A	PPB	15 Feb 2022 15:29:11	U	SMPL
N1535-02MSD - 1	368MSD	16057	3.5886	-7470A	PPB	15 Feb 2022 15:31:26	U	SMPL
PB142692BL - 1	PBW	217	0.0275	-7470A	PPB	15 Feb 2022 15:33:42	U	SMPL
PB142692BS - 1	LCSW	15011	3.3535	-7470A	PPB	15 Feb 2022 15:35:58	U	SMPL
N1515-05 - 1	OW-1	253	0.0356	-7470A	PPB	15 Feb 2022 15:38:14	U	SMPL
N1515-05DUP - 1	OW-1DUP	264	0.0381	-7470A	PPB	15 Feb 2022 15:40:31	U	SMPL
N1515-05MS - 1	OW-1MS	16120	3.6028	-7470A	PPB	15 Feb 2022 15:42:48	U	SMPL
N1515-05MSD - 1	OW-1MSD	14790	3.3038	-7470A	PPB	15 Feb 2022 15:45:06	U	SMPL
PB142693BL - 1	PBW	204	0.0246	-7470A	PPB	15 Feb 2022 15:47:22	U	SMPL
PB142693BS - 1	LCSW	14489	3.2361	-7470A	PPB	15 Feb 2022 15:49:38	U	SMPL
CCV89 - 1	CCV89	21359	4.7806	-7470A	PPB	15 Feb 2022 15:51:54	U	SMPL
CCB89 - 1	CCB89	-12	-0.0239	-7470A	PPB	15 Feb 2022 15:54:09	U	SMPL
N1477-07 - 1	20220065-FIELD-BLANK	194	0.0224	-7470A	PPB	15 Feb 2022 15:56:25	U	SMPL
N1477-08 - 1	20220072-104-21-SITE-WATER	336	0.0543	-7470A	PPB	15 Feb 2022 15:58:41	U	SMPL
N1477-09 - 1	20220073-105-21-SITE-WATER	301	0.0464	-7470A	PPB	15 Feb 2022 16:00:57	U	SMPL
N1513-01 - 1	20220097-FIELD-BLANK-2	189	0.0212	-7470A	PPB	15 Feb 2022 16:03:13	U	SMPL
N1513-02 - 1	20220066-096-21-SITE-WATER	319	0.0505	-7470A	PPB	15 Feb 2022 16:05:29	U	SMPL
N1513-03 - 1	20220067-097-21-SITE-WATER	288	0.0435	-7470A	PPB	15 Feb 2022 16:07:45	U	SMPL
N1513-04 - 1	20220068-098-21-SITE-WATER	254	0.0359	-7470A	PPB	15 Feb 2022 16:10:02	U	SMPL
N1513-05 - 1	20220069-101-21-SITE-WATER	278	0.0413	-7470A	PPB	15 Feb 2022 16:12:19	U	SMPL
N1513-06 - 1	20220070-102-21-SITE-WATER	280	0.0417	-7470A	PPB	15 Feb 2022 16:14:37	U	SMPL
N1513-07 - 1	20220071-103-21-SITE-WATER	251	0.0352	-7470A	PPB	15 Feb 2022 16:16:53	U	SMPL
CCV90 - 1	CCV90	21172	4.7385	-7470A	PPB	15 Feb 2022 16:19:09	U	SMPL
CCB90 - 1	CCB90	-54	-0.0334	-7470A	PPB	15 Feb 2022 16:21:24	U	SMPL
N1513-07DUP - 1	20220071-103-21-SITE-WATERDUP	268	0.0390	-7470A	PPB	15 Feb 2022 16:23:41	U	SMPL
N1513-07MS - 1	20220071-103-21-SITE-WATERMS	16283	3.6394	-7470A	PPB	15 Feb 2022 16:25:56	U	SMPL
N1513-07MSD - 1	20220071-103-21-SITE-WATERMSD	15334	3.4261	-7470A	PPB	15 Feb 2022 16:28:12	U	SMPL
N1515-04 - 1	OW-1	33574	7.5267	-7470A	PPB	15 Feb 2022 16:30:29	U	SMPL
N1530-01 - 1	20220108-FIELD-BLANK	73	-0.0048	-7470A	PPB	15 Feb 2022 16:32:45	U	SMPL
PB142702BL - 1	PBW	192	0.0219	-7470A	PPB	15 Feb 2022 16:35:02	U	SMPL
PB142702BS - 1	LCSW	15028	3.3573	-7470A	PPB	15 Feb 2022 16:37:19	U	SMPL
N1511-03 - 1	SPLP-LEACHATE-AMENDED-NACOTE-1-COMP	-26	-0.0271	-7470A	PPB	15 Feb 2022 16:39:36	U	SMPL
N1511-03DUP - 1	SPLP-LEACHATE-AMENDED-NACOTE-1-COMP DUP	3	-0.0206	-7470A	PPB	15 Feb 2022 16:41:53	U	SMPL
N1511-03MS - 1	SPLP-LEACHATE-AMENDED-NACOTE-1-COMPMS	14663	3.2752	-7470A	PPB	15 Feb 2022 16:44:09	U	SMPL
CCV91 - 1	CCV91	21239	4.7536	-7470A	PPB	15 Feb 2022 16:46:25	U	SMPL
CCB91 - 1	CCB91	-57	-0.0341	-7470A	PPB	15 Feb 2022 16:48:41	U	SMPL
N1511-03MSD - 1	SPLP-LEACHATE-AMENDED-NACOTE-1-COMPMSD	14702	3.2840	-7470A	PPB	15 Feb 2022 16:50:56	U	SMPL
PB142677TB - 1	PB142677TB	176	0.0183	-7470A	PPB	15 Feb 2022 16:53:13	U	SMPL
PB142681TB - 1	PB142681TB	207	0.0253	-7470A	PPB	15 Feb 2022 16:55:29	U	SMPL
PB142702TB - 1	PB142690TB	223	0.0289	-7470A	PPB	15 Feb 2022 16:57:44	U	SMPL
CCV92 - 1	CCV92	20870	4.6707	-7470A	PPB	15 Feb 2022 17:00:00	U	SMPL
CCB92 - 1	CCB92	-29	-0.0278	-7470A	PPB	15 Feb 2022 17:02:17	U	SMPL
N1535-02LX5 - 1		64	-0.0069	-7470A	PPB	15 Feb 2022 17:04:34	U	SMPL
N1535-02A - 1		16613	3.7136	-7470A	PPB	15 Feb 2022 17:06:51	U	SMPL
N1515-05LX5 - 1		28	-0.0149	-7470A	PPB	15 Feb 2022 17:09:09	U	SMPL
N1515-05A - 1		16720	3.7377	-7470A	PPB	15 Feb 2022 17:11:25	U	SMPL
N1513-07LX5 - 1		26	-0.0154	-7470A	PPB	15 Feb 2022 17:13:42	U	SMPL
N1513-07A - 1		16701	3.7334	-7470A	PPB	15 Feb 2022 17:15:58	U	SMPL
N1511-03LX5 - 1		-35	-0.0291	-7470A	PPB	15 Feb 2022 17:18:14	U	SMPL
N1511-03A - 1		18773	4.1992	-7470A	PPB	15 Feb 2022 17:20:30	U	SMPL
CCV93 - 1	CCV93	20848	4.6657	-7470A	PPB	15 Feb 2022 17:22:47	U	SMPL
CCB93 - 1	CCB93	-36	-0.0293	-7470A	PPB	15 Feb 2022 17:25:03	U	SMPL