



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118681

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/15/2022	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	02/15/2022	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	02/15/2022	16:12
4	ICV	1	Water	NA	NA	20.4	7.00	02/15/2022	16:15
5	CCV1	1	Water	NA	NA	20.3	2.01	02/15/2022	16:16
6	N1552-02	1	Solid	20.03	20	20.1	7.97	02/15/2022	16:20
7	N1552-02DUP	1	Solid	20.04	20	20.2	7.98	02/15/2022	16:21
8	N1552-05	1	Solid	20.05	20	20.7	8.66	02/15/2022	16:23
9	CCV2	1	Water	NA	NA	20.4	12.03	02/15/2022	16:25

W3118681

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1552 WorkList ID : 156874 Department : Wet-Chemistry Date : 02-15-2022 15:58:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
02/15/2022	Solid	N1552-02	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-1	02/15/2022	9045D
02/15/2022	Solid	N1552-05	Corrosivity	Cool 4 deg C	PSEG03	M11	TP-2	02/15/2022	9045D

Date/Time 02-15-22 16:00

Raw Sample Received by: YD/60C1

Raw Sample Relinquished by: CP/5M

Date/Time 02-15-22 17:30

Raw Sample Received by: CP/5M

Raw Sample Relinquished by: YD/60C1