



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118683

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/15/2022	16:38
2	CAL2	1	Water	NA	NA	20.3	7.01	02/15/2022	16:39
3	CAL3	1	Water	NA	NA	20.3	10.03	02/15/2022	16:40
4	ICV	1	Water	NA	NA	20.4	6.99	02/15/2022	16:42
5	CCV1	1	Water	NA	NA	20.4	2.01	02/15/2022	16:45
6	N1554-01	1	Solid	20.03	20	20.7	4.76	02/15/2022	16:46
7	N1554-01DUP	1	Solid	20.04	20	20.8	4.77	02/15/2022	16:47
8	CCV2	1	Water	NA	NA	20.4	12.03	02/15/2022	16:50

WORKLIST(Hardcopy Internal Chain)

VB 18683

WorkList Name : PH N1554

WorkList ID : 156876

Department : Wet-Chemistry

Date : 02-15-2022 16:30:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/15/2022	Solid	N1554-01	pH	Cool 4 deg C	PSEG05	M11	BU-02-021522	02/15/2022	9045D

Date/Time 02-15-22 16132
 Raw Sample Received by: JD CSN
 Raw Sample Relinquished by: JD CSN

Date/Time 02-15-22 17140
 Raw Sample Received by: JD CSN
 Raw Sample Relinquished by: JD CSN