



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB118692
Analysis Date: 02/16/2022
BalanceID: WCSC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 02/16/2022
Extraction IN Time: 11:15
Extraction OUT Time: 11:45
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB118692BL	LB118692BL	WATER	1.3	1000	100	3.0537	3.0537	3.01	3.0538	3.0538	0.0001	0.1
2	LB118692BS	LB118692BS	WATER	1.3	1000	100	3.0147	3.0147	3.02	3.0317	3.0317	0.0170	17
3	N1549-01	DRAIN-WATER-TANK-1	WATER	1.6	1000	100	3.0469	3.0469	3.03	3.0598	3.0598	0.0129	12.9
4	N1549-02	N1549-01MS	WATER	1.6	1000	100	3.0856	3.0856	3.04	3.1182	3.1182	0.0326	32.6
5	N1549-03	N1549-01MSD	WATER	1.6	1000	100	3.0253	3.0253	3.04	3.0582	3.0582	0.0329	32.9
6	N1555-01	GRAB	WATER	1.6	1000	100	3.0648	3.0648	3.05	3.0852	3.0852	0.0204	20.4



QC Batch# LB118692

Test: TPH

Analysis Date: 02/16/2022

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2885
pH Paper 0-14	M4909
Sodium Sulfate	EP2222
1:1 HCL	WP92935
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP92936
LCSWD	5.00 ML	WP92937
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 13:01

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 12:15

Bal Check Time: 11:30 Out OVEN TEMP1: 70 °C Dessicator Time Out1 13:30

Out Time1	13:00
-----------	-------

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 14:31

In Time2: 14:00

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 14:00

Out OVEN TEMP2: 71 °C Dessicator Time Out2 15:00

Bal Check Time: 15:30

Out Time2: 14:30