


LB 118693

7471B

INSTRUMENT ID: CV1

 μ Abs.:

50601



Linear

A= 0.0000e+000

B= 1.9766e-004

C= 2.9612e-002

Rho= 0.9998790

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.044	0.044	72	0.000	72					
0.2	0.200	0.232	0.032	1024	0.0 %	1024					16
2.5	2.500	2.464	-0.036	12314	0.0 %	12314					-1
5.0	5.000	4.887	-0.113	24575	0.0 %	24575					-2
7.5	7.500	7.542	0.042	38007	0.0 %	38007					1
10.0	10.000	10.031	0.031	50601	0.0 %	50601					0

LB118693 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Feb 2022 12:28:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.00007471B	PPB		16 Feb 2022 12:32:48	S	Std
0.2 - 1	50.2	1024	-	0.20007471B	PPB		16 Feb 2022 12:35:05	S	Std
2.5 - 1	52.5	12314	-	2.50007471B	PPB		16 Feb 2022 12:37:23	S	Std
5.0 - 1	55	24575	-	5.00007471B	PPB		16 Feb 2022 12:39:41	S	Std
7.5 - 1	57.5	38007	-	7.50007471B	PPB		16 Feb 2022 12:41:58	S	Std
10.0 - 1	510	50601	-	10.00007471B	PPB		16 Feb 2022 12:46:48	S	Std
ICV90 - 1	ICV90	20885	4.1577	-7471B	PPB		16 Feb 2022 12:49:34	U	SMPL
ICB90 - 1	ICB90	54	0.0403	-7471B	PPB		16 Feb 2022 12:51:48	U	SMPL
CCV94 - 1	CCV94	24703	4.9124	-7471B	PPB		16 Feb 2022 12:54:04	U	SMPL
CCB94 - 1	CCB94	57	0.0409	-7471B	PPB		16 Feb 2022 12:56:19	U	SMPL
CRA - 1	CRA	1027	0.2326	-7471B	PPB		16 Feb 2022 12:58:34	U	SMPL
HighStd - 1	HighStd	50433	9.9982	-7471B	PPB		16 Feb 2022 13:00:50	U	SMPL
ChkStd - 1	ChkStd	35599	7.0661	-7471B	PPB		16 Feb 2022 13:03:06	U	SMPL
PB142723BL - 1	PB142723BL	-68	0.0162	-7471B	PPB		16 Feb 2022 13:05:23	U	SMPL
PB142723BS - 1	PB142723BS	19204	3.8255	-7471B	PPB		16 Feb 2022 13:07:40	U	SMPL
N1532-01 - 1	VNJ-209	2078	0.4403	-7471B	PPB		16 Feb 2022 13:09:57	U	SMPL
N1535-01 - 1	368	149542	29.5880	-7471B	PPB		16 Feb 2022 13:12:14	U	SMPL
N1552-02 - 1	TP-1	46457	9.2123	-7471B	PPB		16 Feb 2022 13:14:31	U	SMPL
N1552-05 - 1	TP-2	1926	0.4103	-7471B	PPB		16 Feb 2022 13:16:51	U	SMPL
CCV95 - 1	CCV95	24820	4.9355	-7471B	PPB		16 Feb 2022 13:19:08	U	SMPL
CCB95 - 1	CCB95	53	0.0401	-7471B	PPB		16 Feb 2022 13:21:23	U	SMPL
N1554-01 - 1	BU-03-021522	700	0.1680	-7471B	PPB		16 Feb 2022 13:23:39	U	SMPL
N1554-01DUP - 1	BU-03-021522DUP	512	0.1308	-7471B	PPB		16 Feb 2022 13:25:55	U	SMPL
N1554-01MS - 1	BU-03-021522MS	22676	4.5117	-7471B	PPB		16 Feb 2022 13:28:11	U	SMPL
N1554-01MSD - 1	BU-03-021522MSD	21057	4.1917	-7471B	PPB		16 Feb 2022 13:30:27	U	SMPL
N1554-01LX5 - 1		84	0.0462	-7471B	PPB		16 Feb 2022 13:32:44	U	SMPL
N1554-01A - 1		21958	4.3698	-7471B	PPB		16 Feb 2022 13:35:00	U	SMPL
N1535-01DLX5 - 1		33585	6.6680	-7471B	PPB		16 Feb 2022 13:37:17	U	SMPL
CCV96 - 1	CCV96	24730	4.9177	-7471B	PPB		16 Feb 2022 13:39:34	U	SMPL
CCB96 - 1	CCB96	51	0.0397	-7471B	PPB		16 Feb 2022 13:41:52	U	SMPL