


LB 118702

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

39209



Linear

A= 0.0000e+000

B= 2.5660e-004 *slope*C= 4.5538e-003 *y-intercept*

Rho= 0.9999337

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>Std</i>
0.00	0.000	-0.001	-0.001	-23	0.000	-23					16
0.05	0.050	0.058	0.008	207	0.0 %	207	0				3
0.20	0.200	0.206	0.006	785	0.0 %	785					1
2.50	2.500	2.528	0.028	9835	0.0 %	9835					0
5.00	5.000	4.978	-0.022	19382	0.0 %	19382					0
7.50	7.500	7.416	-0.084	28882	0.0 %	28882					1
10.0	10.000	10.066	0.066	39209	0.0 %	39209					1

LB118702
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-23	-		0 SFAM01.1	PPB	2/16/2022 15:50	Std
	0.05 S01	207	-	0.05	SFAM01.1	PPB	2/16/2022 15:52	Std
	0.2 S02	785	-	0.2	SFAM01.1	PPB	2/16/2022 15:55	Std
	2.5 S03	9835	-	2.5	SFAM01.1	PPB	2/16/2022 15:57	Std
	5 S04	19382	-	5	SFAM01.1	PPB	2/16/2022 15:59	Std
	7.5 S05	28882	-	7.5	SFAM01.1	PPB	2/16/2022 16:02	Std
	10 S06	39209	-	10	SFAM01.1	PPB	2/16/2022 16:11	Std
ICV091	ICV091	14813	3.8056	-	SFAM01.1	PPB	2/16/2022 16:13	SMPL
ICB091	ICB091	-28	-0.0026	-	SFAM01.1	PPB	2/16/2022 16:16	SMPL
CCV097	CCV097	18635	4.7863	-	SFAM01.1	PPB	2/16/2022 16:18	SMPL
CCB097	CCB097	-22	-0.0011	-	SFAM01.1	PPB	2/16/2022 16:20	SMPL
PB142726BL	PBW726	36	0.0138	-	SFAM01.1	PPB	2/16/2022 16:22	SMPL
N1331-01	MDL-WATER-QT1	205	0.0572	-	SFAM01.1	PPB	2/16/2022 16:25	SMPL
CCV098	CCV098	18557	4.7663	-	SFAM01.1	PPB	2/16/2022 16:27	SMPL
CCB098	CCB098	-15	0.0007	-	SFAM01.1	PPB	2/16/2022 16:29	SMPL