



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118705

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/16/2022	16:10
2	CAL2	1	Water	NA	NA	20.4	7.00	02/16/2022	16:11
3	CAL3	1	Water	NA	NA	20.4	10.03	02/16/2022	16:12
4	ICV	1	Water	NA	NA	20.5	6.99	02/16/2022	16:15
5	CCV1	1	Water	NA	NA	20.4	2.01	02/16/2022	16:19
6	N1534-01	1	Solid	20.03	20	20.9	7.63	02/16/2022	16:20
7	N1556-02	1	Solid	20.02	20	20.1	10.87	02/16/2022	16:24
8	N1561-03	1	Solid	20.03	20	20.2	7.50	02/16/2022	16:26
9	N1561-06	1	Solid	20.05	20	20.2	8.36	02/16/2022	16:29
10	N1564-02	1	Solid	20.04	20	20.5	7.97	02/16/2022	16:33
11	N1564-05	1	Solid	20.02	20	20.3	8.30	02/16/2022	16:36
12	N1571-02	1	Solid	20.01	20	20.6	7.83	02/16/2022	16:40
13	N1571-02DUP	1	Solid	20.02	20	20.7	7.84	02/16/2022	16:41
14	CCV2	1	Water	NA	NA	20.4	12.03	02/16/2022	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1556 WorkList ID : 156897 Department : Wet-Chemistry Date : 02-16-2022 12:29:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/11/2022	Solid	N1534-01	Corrosivity	Cool 4 deg C	ENTA05	M11	NAPL-SOCK	02/11/2022	9045D
02/16/2022	Solid	N1556-02	Corrosivity	Cool 4 deg C	ENTA05	N11	N3-SURFICIAL-SOIL	02/16/2022	9045D
02/16/2022	Solid	N1561-03	Corrosivity	Cool 4 deg C	TULL02	M11	TP-1-COMP	02/16/2022	9045D
02/16/2022	Solid	N1561-06	Corrosivity	Cool 4 deg C	TULL02	M11	TP-2-COMP	02/16/2022	9045D
02/16/2022	Solid	N1564-02	Corrosivity	Cool 4 deg C	PSEG03	N11	TP-2	02/16/2022	9045D
02/16/2022	Solid	N1564-05	Corrosivity	Cool 4 deg C	PSEG03	N11	TP-3	02/16/2022	9045D
02/16/2022	Solid	N1571-02	Corrosivity	Cool 4 deg C	PSEG03	N11	TP-3	02/16/2022	9045D

Date/Time 02.16.22 16:00
 Raw Sample Received by: JP (600)
 Raw Sample Relinquished by: CP SM

Date/Time 02.16.22 17:20
 Raw Sample Received by: SP SM
 Raw Sample Relinquished by: JP (600)