



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118706

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/16/2022	17:25
2	CAL2	1	Water	NA	NA	20.3	7.00	02/16/2022	17:26
3	CAL3	1	Water	NA	NA	20.3	10.03	02/16/2022	17:28
4	ICV	1	Water	NA	NA	20.4	6.99	02/16/2022	17:30
5	CCV1	1	Water	NA	NA	20.4	2.01	02/16/2022	17:32
6	N1577-02	1	Water	NA	NA	20.3	5.92	02/16/2022	17:35
7	N1577-02DUP	1	Water	NA	NA	20.4	5.93	02/16/2022	17:36
8	N1577-03	1	Water	NA	NA	20.3	6.45	02/16/2022	17:39
9	N1577-04	1	Water	NA	NA	20.1	6.01	02/16/2022	17:40
10	N1577-05	1	Water	NA	NA	20.1	5.65	02/16/2022	17:42
11	N1577-06	1	Water	NA	NA	20.3	6.08	02/16/2022	17:45
12	CCV2	1	Water	NA	NA	20.6	12.02	02/16/2022	17:46

WORKLIST(Hardcopy Internal Chain)

18118106

WorkList Name : ph n1577 WorkList ID : 156911 Department : Wet-Chemistry Date : 02-16-2022 17:15:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/16/2022	Water	N1577-02	pH	Cool 4 deg C	PSEG03	N11	1741	02/16/2022	9040C
02/16/2022	Water	N1577-03	pH	Cool 4 deg C	PSEG03	N11	1742	02/16/2022	9040C
02/16/2022	Water	N1577-04	pH	Cool 4 deg C	PSEG03	N11	1743	02/16/2022	9040C
02/16/2022	Water	N1577-05	pH	Cool 4 deg C	PSEG03	N11	1744	02/16/2022	9040C
02/16/2022	Water	N1577-06	pH	Cool 4 deg C	PSEG03	N11	1745	02/16/2022	9040C

Date/Time 02/16/22 17:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02-16-22 18:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]