


LB118710

SFAM01.1

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.7524e-004 slope

C= 3.4691e-003 y-intercept

Rho= 0.9999422

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	22	0.000	22					1
0.05	0.050	0.063	0.013	217	0.0 %	217	0				26
0.20	0.200	0.212	0.012	759	0.0 %	759					6
2.50	2.500	2.483	-0.017	9007	0.0 %	9007					1
5.00	5.000	4.999	-0.001	18149	0.0 %	18149					0
7.50	7.500	7.420	-0.080	26946	0.0 %	26946					1
10.0	10.000	10.063	0.063	36550	0.0 %	36550					1

LB118710
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	22	-		0 SFAM01.1	PPB	2/17/2022 9:43	Std
	0.05 S01	217	-		0.05 SFAM01.1	PPB	2/17/2022 9:45	Std
	0.2 S02	759	-		0.2 SFAM01.1	PPB	2/17/2022 9:47	Std
	2.5 S03	9007	-		2.5 SFAM01.1	PPB	2/17/2022 9:52	Std
	5 S04	18149	-		5 SFAM01.1	PPB	2/17/2022 9:54	Std
	7.5 S05	26946	-		7.5 SFAM01.1	PPB	2/17/2022 9:56	Std
	10 S06	36550	-		10 SFAM01.1	PPB	2/17/2022 10:01	Std
ICV092	ICV092	14735	4.0591 -		SFAM01.1	PPB	2/17/2022 10:04	SMPL
ICB092	ICB092	-28	-0.0042 -		SFAM01.1	PPB	2/17/2022 10:06	SMPL
CCV001	CCV001	18175	5.006 -		SFAM01.1	PPB	2/17/2022 10:08	SMPL
CCB001	CCB001	-32	-0.0053 -		SFAM01.1	PPB	2/17/2022 10:11	SMPL
PB142746BL	PBW746	14	0.0073 -		SFAM01.1	PPB	2/17/2022 10:13	SMPL
N1331-02	MDL-WATER-QT1	194	0.0569 -		SFAM01.1	PPB	2/17/2022 10:15	SMPL
CCV002	CCV002	18124	4.9919 -		SFAM01.1	PPB	2/17/2022 10:18	SMPL
CCB002	CCB002	-25	-0.0034 -		SFAM01.1	PPB	2/17/2022 10:20	SMPL