



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB118711
Analysis Date: 02/17/2022
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 02/17/2022
Extraction IN Time: 09:30
Extraction OUT Time: 10:00
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB118711BL	LB118711BL	WATER	1.3	1000	100	2.9631	2.9631	3.02	2.9632	2.9632	0.0001	0.1
2	LB118711BS	LB118711BS	WATER	1.3	1000	100	2.7438	2.7438	3.01	2.7608	2.7608	0.0170	17
3	LB118711BSD	LB118711BSD	WATER	1.3	1000	100	3.0247	3.0247	3.03	3.0418	3.0418	0.0171	17.1
4	N1577-01	COMP-1741-1745	WATER	1.3	1000	100	2.9189	2.9189	3.04	2.9193	2.9193	0.0004	0.4

VB 118711

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH N1577 WorkList ID : 156928 Department : Wet-Chemistry Date : 02-17-2022 09:20:28

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
02/21/2022	Water	N1577-01	TPH <i>CA</i>	Conc H2SO4 to	PSEG03	N11	COMP-1741-1745	02/16/2022 1664A

Date/Time 02-17-22 09:22
Raw Sample Received by: JD (QC)
Raw Sample Relinquished by: JD (SM)

Date/Time 02-17-22 11:40
Raw Sample Received by: JR (SM)
Raw Sample Relinquished by: JR (QC)