


LB118715

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4661e-004 *slope*

C= -1.6064e-002 *y-intercept*

Rho= 0.9997685

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.015	-0.015	4	0.000	4					
0.05	0.050	0.035	-0.015	207	0.0 %	207	0				-30
0.20	0.200	0.197	-0.003	863	0.0 %	863					-2
2.50	2.500	2.581	0.081	10533	0.0 %	10533					3
5.00	5.000	5.038	0.038	20496	0.0 %	20496					1
7.50	7.500	7.330	-0.170	29787	0.0 %	29787					-2
10.0	10.000	10.084	0.084	40956	0.0 %	40956					1

LB118715
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	4	-		0 SFAM01.1	PPB	2/17/2022 12:30	Std
	0.05 S01	207	-	0.05	SFAM01.1	PPB	2/17/2022 12:32	Std
	0.2 S02	863	-	0.2	SFAM01.1	PPB	2/17/2022 12:35	Std
	2.5 S03	10533	-	2.5	SFAM01.1	PPB	2/17/2022 12:37	Std
	5 S04	20496	-	5	SFAM01.1	PPB	2/17/2022 12:39	Std
	7.5 S05	29787	-	7.5	SFAM01.1	PPB	2/17/2022 12:47	Std
	10 S06	40956	-	10	SFAM01.1	PPB	2/17/2022 12:54	Std
ICV093	ICV093	16003	3.9304 -		SFAM01.1	PPB	2/17/2022 13:00	SMPL
ICB093	ICB093	3	-0.0153 -		SFAM01.1	PPB	2/17/2022 13:02	SMPL
CCV003	CCV003	18955	4.6584 -		SFAM01.1	PPB	2/17/2022 13:04	SMPL
CCB003	CCB003	-4	-0.0171 -		SFAM01.1	PPB	2/17/2022 13:07	SMPL
PB142747BL	PBS747	26	-0.0097 -		SFAM01.1	PPB	2/17/2022 13:09	SMPL
N1331-03	MDL-SOIL-QT1	226	0.0397 -		SFAM01.1	PPB	2/17/2022 13:11	SMPL
CCV004	CCV004	19673	4.8354 -		SFAM01.1	PPB	2/17/2022 13:18	SMPL
CCB004	CCB004	-14	-0.0195 -		SFAM01.1	PPB	2/17/2022 13:21	SMPL