


LB118729

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6724e-004 *slope*C= -9.5535e-003 *y-intercept*

Rho= 0.9999036

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	13	0.000	13					
0.05	0.050	0.040	-0.010	185	0.0 %	185	0				-20
0.20	0.200	0.201	0.001	787	0.0 %	787					1
2.50	2.500	2.536	0.036	9527	0.0 %	9527					1
5.00	5.000	4.930	-0.070	18485	0.0 %	18485					1
7.50	7.500	7.599	0.099	28470	0.0 %	28470					1
10.0	10.000	9.950	-0.050	37267	0.0 %	37267					1

LB118727

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	13	-		0 SFAM01.1	PPB	2/18/2022 10:17	Std
	0.05 S01	185	-		0.05 SFAM01.1	PPB	2/18/2022 10:20	Std
	0.2 S02	787	-		0.2 SFAM01.1	PPB	2/18/2022 10:22	Std
	2.5 S03	9527	-		2.5 SFAM01.1	PPB	2/18/2022 10:24	Std
	5 S04	18485	-		5 SFAM01.1	PPB	2/18/2022 10:26	Std
	7.5 S05	28470	-		7.5 SFAM01.1	PPB	2/18/2022 10:29	Std
	10 S06	37267	-		10 SFAM01.1	PPB	2/18/2022 10:33	Std
ICV095	ICV095	15218	4.0573 -		SFAM01.1	PPB	2/18/2022 10:36	SMPL
ICB095	ICB095	-33	-0.0184 -		SFAM01.1	PPB	2/18/2022 10:39	SMPL
CCV009	CCV009	18214	4.858 -		SFAM01.1	PPB	2/18/2022 10:41	SMPL
CCB009	CCB009	-29	-0.0173 -		SFAM01.1	PPB	2/18/2022 10:43	SMPL
PB142785BL	PBS785	3	-0.0088 -		SFAM01.1	PPB	2/18/2022 10:45	SMPL
N1331-04	MDL-SOIL-QT1	206	0.0455 -		SFAM01.1	PPB	2/18/2022 10:48	SMPL
CCV010	CCV010	18315	4.885 -		SFAM01.1	PPB	2/18/2022 10:50	SMPL
CCB010	CCB010	-25	-0.0162 -		SFAM01.1	PPB	2/18/2022 10:52	SMPL