


LB118730

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3007e-004

C= -8.9916e-003

Rho= 0.9998466

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.0	0.000	-0.008	-0.008	5	0.000	5					1
0.2	0.200	0.197	-0.003	894	0.0 %	894					-2
2.5	2.500	2.469	-0.031	10769	0.0 %	10769					-1
5.0	5.000	5.013	0.013	21827	0.0 %	21827					0
7.5	7.500	7.623	0.123	33173	0.0 %	33173					2
10.0	10.000	9.907	-0.093	43099	0.0 %	43099					-1

LB118730 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Feb 2022 11:27:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	7470A	PPB	18 Feb 2022 11:36:54	S	Std
0.2 - 1	50.2	894	-	0.2000	7470A	PPB	18 Feb 2022 11:39:11	S	Std
2.5 - 1	52.5	10769	-	2.5000	7470A	PPB	18 Feb 2022 11:41:28	S	Std
5.0 - 1	55	21827	-	5.0000	7470A	PPB	18 Feb 2022 11:43:46	S	Std
7.5 - 1	57.5	33173	-	7.5000	7470A	PPB	18 Feb 2022 11:46:03	S	Std
10.0 - 1	510	43099	-	10.0000	7470A	PPB	18 Feb 2022 11:51:07	S	Std
ICV96 - 1	ICV96	17537	4.0257	-7470A	PPB	18 Feb 2022 11:53:52	U	SMPL	
ICB96 - 1	ICB96	-57	-0.0221	-7470A	PPB	18 Feb 2022 11:56:07	U	SMPL	
CCV11 - 1	CCV11	21500	4.9375	-7470A	PPB	18 Feb 2022 12:00:04	U	SMPL	
CCB11 - 1	CCB11	-34	-0.0168	-7470A	PPB	18 Feb 2022 12:02:19	U	SMPL	
CRA - 1	CRA	872	0.1916	-7470A	PPB	18 Feb 2022 12:04:34	U	SMPL	
HighStd - 1	HighStd	43679	10.0402	-7470A	PPB	18 Feb 2022 12:06:50	U	SMPL	
ChkStd - 1	ChkStd	30186	6.9359	-7470A	PPB	18 Feb 2022 12:09:05	U	SMPL	
PB142752BL - 1	PBW	-72	-0.0256	-7470A	PPB	18 Feb 2022 12:11:21	U	SMPL	
PB142752BS - 1	LCSW	15513	3.5601	-7470A	PPB	18 Feb 2022 12:13:37	U	SMPL	
N1455-13 - 1	TP-2	95	0.0129	-7470A	PPB	18 Feb 2022 12:15:54	U	SMPL	
N1455-14 - 1	TP-5	41	0.0004	-7470A	PPB	18 Feb 2022 12:18:11	U	SMPL	
N1534-01 - 1	NAPL-SOCK	20	-0.0044	-7470A	PPB	18 Feb 2022 12:20:28	U	SMPL	
N1556-02 - 1	NE-SURFICIAL-SOIL	2187	0.4942	-7470A	PPB	18 Feb 2022 12:22:46	U	SMPL	
N1561-03 - 1	TP-1-1-COMP	17	-0.0051	-7470A	PPB	18 Feb 2022 12:25:02	U	SMPL	
CCV12 - 1	CCV12	21085	4.8420	-7470A	PPB	18 Feb 2022 12:27:17	U	SMPL	
CCB12 - 1	CCB12	-31	-0.0161	-7470A	PPB	18 Feb 2022 12:29:33	U	SMPL	
N1561-06 - 1	TP-2-COMP	14	-0.0058	-7470A	PPB	18 Feb 2022 12:31:48	U	SMPL	
N1561-06DUP - 1	TP-2-COMPDUP	15	-0.0055	-7470A	PPB	18 Feb 2022 12:34:04	U	SMPL	
N1561-06MS - 1	TP-2-COMPMS	15832	3.6335	-7470A	PPB	18 Feb 2022 12:36:20	U	SMPL	
N1561-06MSD - 1	TP-2-COMPMSD	15680	3.5985	-7470A	PPB	18 Feb 2022 12:41:53	U	SMPL	
PB142753BL - 1	PBW	-15	-0.0124	-7470A	PPB	18 Feb 2022 12:44:09	U	SMPL	
PB142753BS - 1	LCSW	13957	3.2021	-7470A	PPB	18 Feb 2022 12:46:25	U	SMPL	
N1577-01 - 1	COMP-1741-1745	-1	-0.0092	-7470A	PPB	18 Feb 2022 12:48:42	U	SMPL	
N1577-01DUP - 1	COMP-1741-1745DUP	23	-0.0037	-7470A	PPB	18 Feb 2022 12:51:00	U	SMPL	
N1577-01MS - 1	COMP-1741-1745MS	15791	3.6240	-7470A	PPB	18 Feb 2022 12:53:17	U	SMPL	
N1577-01MSD - 1	COMP-1741-1745MSD	16066	3.6873	-7470A	PPB	18 Feb 2022 12:55:32	U	SMPL	
CCV13 - 1	CCV13	21062	4.8367	-7470A	PPB	18 Feb 2022 12:57:48	U	SMPL	
CCB13 - 1	CCB13	-62	-0.0233	-7470A	PPB	18 Feb 2022 13:00:04	U	SMPL	
PB142754BL - 1	PBW	7	-0.0074	-7470A	PPB	18 Feb 2022 13:02:20	U	SMPL	
PB142754BS - 1	LCSW	14208	3.2598	-7470A	PPB	18 Feb 2022 13:04:36	U	SMPL	
N1481-01 - 1	RW7B-GW-20220209	-31	-0.0161	-7470A	PPB	18 Feb 2022 13:06:51	U	SMPL	
N1481-01DUP - 1	RW7B-GW-20220209DUP	4	-0.0081	-7470A	PPB	18 Feb 2022 13:09:08	U	SMPL	
N1481-01MS - 1	RW7B-GW-20220209MS	17222	3.9533	-7470A	PPB	18 Feb 2022 13:14:43	U	SMPL	
N1481-01MSD - 1	RW7B-GW-20220209MSD	16234	3.7259	-7470A	PPB	18 Feb 2022 13:16:59	U	SMPL	
N1581-27 - 1	FB02-20220215	-51	-0.0207	-7470A	PPB	18 Feb 2022 13:19:16	U	SMPL	
PB142724TB - 1	PB142724TB	6	-0.0076	-7470A	PPB	18 Feb 2022 13:21:33	U	SMPL	
PB142731TB - 1	PB142731TB	12	-0.0062	-7470A	PPB	18 Feb 2022 13:23:50	U	SMPL	
CCV14 - 1	CCV14	21134	4.8533	-7470A	PPB	18 Feb 2022 13:26:06	U	SMPL	
CCB14 - 1	CCB14	-41	-0.0184	-7470A	PPB	18 Feb 2022 13:28:22	U	SMPL	
N1561-06LX5 - 1		-10	-0.0113	-7470A	PPB	18 Feb 2022 13:30:39	U	SMPL	
N1561-06A - 1		17442	4.0039	-7470A	PPB	18 Feb 2022 13:32:54	U	SMPL	
N1577-01LX5 - 1		-53	-0.0212	-7470A	PPB	18 Feb 2022 13:35:10	U	SMPL	
N1577-01A - 1		17266	3.9634	-7470A	PPB	18 Feb 2022 13:37:27	U	SMPL	
N1481-01LX5 - 1		-49	-0.0203	-7470A	PPB	18 Feb 2022 13:39:42	U	SMPL	
N1481-01A - 1		14895	3.4179	-7470A	PPB	18 Feb 2022 13:41:58	U	SMPL	
CCV15 - 1	CCV15	21335	4.8995	-7470A	PPB	18 Feb 2022 13:44:15	U	SMPL	
CCB15 - 1	CCB15	-48	-0.0200	-7470A	PPB	18 Feb 2022 13:46:31	U	SMPL	