



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118732

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	02/18/2022	11:32
2	CAL2	1	Water	NA	NA	20.8	7.01	02/18/2022	11:33
3	CAL3	1	Water	NA	NA	20.7	10.03	02/18/2022	11:35
4	ICV	1	Water	NA	NA	20.7	6.98	02/18/2022	11:40
5	CCV1	1	Water	NA	NA	20.8	2.01	02/18/2022	11:42
6	N1595-01	1	Solid	20.03	20	20.2	6.80	02/18/2022	11:45
7	N1595-02	1	Solid	20.04	20	20.8	8.41	02/18/2022	11:46
8	N1595-02DUP	1	Solid	20.03	20	20.9	8.42	02/18/2022	11:47
9	CCV2	1	Water	NA	NA	20.6	12.03	02/18/2022	11:50

WORKLIST(Hardcopy Internal Chain)

78118732

WorkList Name : corrosivity n1595 s WorkList ID : 156969 Department : Wet-Chemistry Date : 02-18-2022 11:31:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/17/2022	Solid	N1595-01	Corrosivity	Cool 4 deg C	PSEG03	N11	COMP-1	02/17/2022	9045D
02/17/2022	Solid	N1595-02	Corrosivity	Cool 4 deg C	PSEG03	N11	5552	02/17/2022	9045D

02-18-22 1135
 Date/Time 02-18-22 1135
 Raw Sample Received by: JH (WC)
 Raw Sample Relinquished by: JH (WC)

02-18-22 13100
 Date/Time 02-18-22 13100
 Raw Sample Received by: JH (WC)
 Raw Sample Relinquished by: JH (WC)