



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB118743

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP95377

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB118743BL		1	21.4	0.88	0.94	02/18/2022	15:20
2	LB118743BS	1413	1	21.6	1334.00	1426.65	02/18/2022	15:23
3	N1613-01		1	20.3	1.13	1.24	02/18/2022	15:27
4	N1613-01DUP		1	20.4	1.14	1.25	02/18/2022	15:32

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$

WORKLIST(Hardcopy Internal Chain)

118793

WorkList Name : Conductance-021822 WorkList ID : 156987 Department : Wet-Chemistry Date : 02-18-2022 13:59:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/04/2022	Water	N1613-01	Conductance	Cool 4 deg C	STER06	N33	POOL-WATER	02/15/2022	SM2510 B

Date/Time 02/18/22 1:20
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 02/18/22 3:40
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)