



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118745

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	02/18/2022	09:45
2	N1577-01		1	10	100.00	>212.0	>212.0	02/18/2022	10:10
3	N1577-01DUP		1	10	100.00	>212.0	>212.0	02/18/2022	10:35
4	N1595-03		1	15	100.00	>212.0	>212.0	02/18/2022	10:58
5	N1595-04		1	14	100.00	>212.0	>212.0	02/18/2022	11:25
6	N1595-05		1	12	100.00	>212.0	>212.0	02/18/2022	12:45
7	N1595-06		1	13	100.00	>212.0	>212.0	02/18/2022	13:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp21822 WorkList ID : 156956 Department : Wet-Chemistry Date : 02-18-2022 09:17:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
02/21/2022	Water	N1577-01	Flash Point	Cool 4 deg C	PSEG03	N11	COMP-1741-1745	02/16/2022	1010B
02/22/2022	Water	N1595-03	Flash Point	Cool 4 deg C	PSEG03	N11	5342	02/17/2022	1010B
02/22/2022	Water	N1595-04	Flash Point	Cool 4 deg C	PSEG03	N11	5326	02/17/2022	1010B
02/22/2022	Water	N1595-05	Flash Point	Cool 4 deg C	PSEG03	N11	5308	02/17/2022	1010B
02/22/2022	Water	N1595-06	Flash Point	Cool 4 deg C	PSEG03	N11	5286	02/17/2022	1010B

Date/Time 02/18/22 09:25
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: RJCET-1667

Date/Time 02/18/22 13:15
 Raw Sample Received by: RJCET-1667
 Raw Sample Relinquished by: KS(WC)