



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118749

Slope : 98.7

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/18/2022	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	02/18/2022	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	02/18/2022	16:17
4	ICV	1	Water	NA	NA	20.7	6.99	02/18/2022	16:19
5	CCV1	1	Water	NA	NA	20.4	2.01	02/18/2022	16:20
6	N1611-05	1	Water	NA	NA	20.2	11.68	02/18/2022	16:22
7	N1611-06	1	Water	NA	NA	20.3	5.75	02/18/2022	16:28
8	N1611-07	1	Water	NA	NA	20.5	6.51	02/18/2022	16:30
9	N1613-01	1	Water	NA	NA	20.3	4.95	02/18/2022	16:33
10	N1613-01DUP	1	Water	NA	NA	20.4	4.96	02/18/2022	16:34
11	CCV2	1	Water	NA	NA	20.5	12.03	02/18/2022	16:35

WORKLIST(Hardcopy Internal Chain)

WB 11849

WorkList Name : ph n1613

WorkList ID : 156991

Department : Wet-Chemistry

Date : 02-18-2022 14:07:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/18/2022	Water	N1611-05	pH	Cool 4 deg C	PSEG03	M11	1746-WATER	02/18/2022	9040C
02/18/2022	Water	N1611-06	pH	Cool 4 deg C	PSEG03	M11	1749	02/18/2022	9040C
02/18/2022	Water	N1611-07	pH	Cool 4 deg C	PSEG03	M11	1070	02/18/2022	9040C
02/15/2022	Water	N1613-01	pH	Cool 4 deg C	STER06	N33	POOL-WATER	02/15/2022	9040C

16:10

Date/Time 02.18.22 15:45
 Raw Sample Received by: J0160C
 Raw Sample Relinquished by: Cpl Sm

17:30

Date/Time 02.18.22
 Raw Sample Received by: Cpl Sm
 Raw Sample Relinquished by: J0160C