



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118750

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/18/2022	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	02/18/2022	16:16
3	CAL3	1	Water	NA	NA	20.3	10.03	02/18/2022	16:17
4	ICV	1	Water	NA	NA	20.7	6.99	02/18/2022	16:19
5	CCV1	1	Water	NA	NA	20.4	2.01	02/18/2022	16:20
6	N1611-01	1	Solid	20.03	20	20.1	11.46	02/18/2022	16:40
7	N1611-02	1	Solid	20.04	20	20.4	11.50	02/18/2022	16:42
8	N1611-03	1	Solid	20.01	20	20.2	11.61	02/18/2022	16:45
9	N1611-03DUP	1	Solid	20.01	20	20.3	11.62	02/18/2022	16:46
10	CCV2	1	Water	NA	NA	20.5	12.03	02/18/2022	16:50

118150

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHN1614 S WorkList ID : 156996 Department : Wet-Chemistry Date : 02-18-2022 16:16:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/18/2022	Solid	N1611-01	pH	Cool 4 deg C	PSEG03	M11	1746-SOIL	02/18/2022	9045D
02/18/2022	Solid	N1611-02	pH	Cool 4 deg C	PSEG03	M11	1747	02/18/2022	9045D
02/18/2022	Solid	N1611-03	pH	Cool 4 deg C	PSEG03	M11	1748	02/18/2022	9045D

Date/Time 02.18.22 15:40
 Raw Sample Received by: Op Sam
 Raw Sample Relinquished by: Op Sam

Date/Time 02.18.22 17:30
 Raw Sample Received by: Op Sam
 Raw Sample Relinquished by: Op Sam