

LB118762


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

38083



A= 0.0000e+000

B= 2.6407e-004

C= -1.9808e-002

Rho= 0.9999055

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.012	-0.012	28	0.000	28					
0.2	0.200	0.195	-0.005	812	0.0 %	812					
2.5	2.500	2.516	0.016	9604	0.0 %	9604					
5.0	5.000	5.063	0.063	19249	0.0 %	19249					
7.5	7.500	7.402	-0.098	28105	0.0 %	28105					
10.0	10.000	10.037	0.037	38083	0.0 %	38083					

1
 -3
 1
 1
 0

LB118762 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Feb 2022 08:51:10

Sample ID	Extended ID	μ Abs.	Conc.	Std. Cond	Method	Units	Date	Type	Type
0.0 - 1	502	28	-	0.00007470A		PPB	21 Feb 2022 08:57:57	S	Std
0.2 - 1	502	812	-	0.20007470A		PPB	21 Feb 2022 09:00:16	S	Std
2.5 - 1	52.5	9604	-	2.50007470A		PPB	21 Feb 2022 09:02:33	S	Std
5.0 - 1	55	19249	-	5.00007470A		PPB	21 Feb 2022 09:04:51	S	Std
7.5 - 1	57.5	28105	-	7.50007470A		PPB	21 Feb 2022 09:07:08	S	Std
10.0 - 1	510	38083	-	10.00007470A		PPB	21 Feb 2022 09:11:57	S	Std
ICV98 - 1	ICV98	14422	3.7885	-7470A		PPB	21 Feb 2022 09:14:40	U	SMPL
ICB98 - 1	ICB98	-27	-0.0269	-7470A		PPB	21 Feb 2022 09:16:55	U	SMPL
CCV19 - 1	CCV19	18922	4.9768	-7470A		PPB	21 Feb 2022 09:19:11	U	SMPL
CCB19 - 1	CCB19	-33	-0.0285	-7470A		PPB	21 Feb 2022 09:21:26	U	SMPL
CRA - 1	CRA	773	0.1843	-7470A		PPB	21 Feb 2022 09:23:41	U	SMPL
HighStd - 1	HighStd	37694	9.9339	-7470A		PPB	21 Feb 2022 09:25:57	U	SMPL
ChkStd - 1	ChkStd	23819	6.2700	-7470A		PPB	21 Feb 2022 09:28:13	U	SMPL
PB142820BL - 1	PBW	-31	-0.0280	-7470A		PPB	21 Feb 2022 09:33:05	U	SMPL
PB142820BS - 1	LCSW	15424	4.0531	-7470A		PPB	21 Feb 2022 09:35:22	U	SMPL
N1595-01 - 1	COMP-1	-19	-0.0248	-7470A		PPB	21 Feb 2022 09:37:40	U	SMPL
N1595-02 - 1	5552	7	-0.0180	-7470A		PPB	21 Feb 2022 09:39:57	U	SMPL
N1595-02DUP - 1	5552DUP	13	-0.0164	-7470A		PPB	21 Feb 2022 09:42:15	U	SMPL
N1595-02MS - 1	5552MS	18205	4.7875	-7470A		PPB	21 Feb 2022 09:44:32	U	SMPL
N1595-02MSD - 1	5552MSD	18078	4.7540	-7470A		PPB	21 Feb 2022 09:46:47	U	SMPL
CCV20 - 1	CCV20	19341	5.0875	-7470A		PPB	21 Feb 2022 09:49:03	U	SMPL
CCB20 - 1	CCB20	-35	-0.0291	-7470A		PPB	21 Feb 2022 09:51:19	U	SMPL
PB142821BL - 1	PBW	-2	-0.0203	-7470A		PPB	21 Feb 2022 09:53:35	U	SMPL
PB142821BS - 1	LCSW	17141	4.5065	-7470A		PPB	21 Feb 2022 09:55:50	U	SMPL
N1595-03 - 1	5342	-24	-0.0261	-7470A		PPB	21 Feb 2022 09:58:06	U	SMPL
N1595-03DUP - 1	5342DUP	6	-0.0182	-7470A		PPB	21 Feb 2022 10:00:22	U	SMPL
N1595-03MS - 1	5342MS	15216	3.9982	-7470A		PPB	21 Feb 2022 10:02:38	U	SMPL
N1595-03MSD - 1	5342MSD	16672	4.3827	-7470A		PPB	21 Feb 2022 10:04:55	U	SMPL
N1595-04 - 1	5326	-5	-0.0211	-7470A		PPB	21 Feb 2022 10:07:12	U	SMPL
N1595-05 - 1	5308	-1	-0.0201	-7470A		PPB	21 Feb 2022 10:09:29	U	SMPL
N1595-06 - 1	5286	7	-0.0180	-7470A		PPB	21 Feb 2022 10:11:46	U	SMPL
PB142822BL - 1	PBW	13	-0.0164	-7470A		PPB	21 Feb 2022 10:14:02	U	SMPL
CCV21 - 1	CCV21	18948	4.9837	-7470A		PPB	21 Feb 2022 10:16:18	U	SMPL
CCB21 - 1	CCB21	-42	-0.0309	-7470A		PPB	21 Feb 2022 10:18:34	U	SMPL
PB142822BS - 1	LCSW	14688	3.8588	-7470A		PPB	21 Feb 2022 10:20:50	U	SMPL
N1602-11 - 1	SPLP-LEACHATE-AMENDED-2-COMP	-13	-0.0232	-7470A		PPB	21 Feb 2022 10:23:06	U	SMPL
N1602-12 - 1	SPLP-LEACHATE-AMENDED-3-COMP	11	-0.0169	-7470A		PPB	21 Feb 2022 10:25:22	U	SMPL
N1602-13 - 1	SPLP-LEACHATE-AMENDED-4A-COMP	18	-0.0151	-7470A		PPB	21 Feb 2022 10:27:38	U	SMPL
N1602-14 - 1	SPLP-LEACHATE-AMENDED-4B-COMP	15	-0.0158	-7470A		PPB	21 Feb 2022 10:29:54	U	SMPL
N1602-15 - 1	SPLP-LEACHATE-AMENDED-6-COMP	28	-0.0124	-7470A		PPB	21 Feb 2022 10:32:11	U	SMPL
N1602-15DUP - 1	SPLP-LEACHATE-AMENDED-6-COMPDUP	21	-0.0143	-7470A		PPB	21 Feb 2022 10:34:28	U	SMPL
N1602-15MS - 1	SPLP-LEACHATE-AMENDED-6-COMPMS	15493	4.0714	-7470A		PPB	21 Feb 2022 10:36:45	U	SMPL
N1602-15MSD - 1	SPLP-LEACHATE-AMENDED-6-COMPMSD	15579	4.0941	-7470A		PPB	21 Feb 2022 10:39:03	U	SMPL
CCV22 - 1	CCV22	19369	5.0949	-7470A		PPB	21 Feb 2022 10:41:19	U	SMPL
CCB22 - 1	CCB22	-35	-0.0291	-7470A		PPB	21 Feb 2022 10:43:36	U	SMPL
PB142789TB - 1		15	-0.0158	-7470A		PPB	21 Feb 2022 10:45:52	U	SMPL
PB142790TB - 1		13	-0.0164	-7470A		PPB	21 Feb 2022 10:48:07	U	SMPL
PB142822TB - 1		33	-0.0111	-7470A		PPB	21 Feb 2022 10:50:23	U	SMPL
N1595-02LX5 - 1		5	-0.0185	-7470A		PPB	21 Feb 2022 10:52:40	U	SMPL
N1595-02A - 1		15523	4.0793	-7470A		PPB	21 Feb 2022 10:54:56	U	SMPL
N1595-03LX5 - 1		-19	-0.0248	-7470A		PPB	21 Feb 2022 10:57:12	U	SMPL
N1595-03A - 1		15625	4.1062	-7470A		PPB	21 Feb 2022 10:59:29	U	SMPL
N1602-15LX5 - 1		-15	-0.0238	-7470A		PPB	21 Feb 2022 11:01:46	U	SMPL
N1602-15A - 1		15455	4.0613	-7470A		PPB	21 Feb 2022 11:04:04	U	SMPL
CCV23 - 1	CCV23	19006	4.9990	-7470A		PPB	21 Feb 2022 11:06:21	U	SMPL
CCB23 - 1	CCB23	-21	-0.0254	-7470A		PPB	21 Feb 2022 11:08:37	U	SMPL

→ PB142789TB
 → PB142790TB
 → PB142822TB
 → PB142822TB
 → PB142822TB