

LB118767


7471B

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

45021



A= 0.0000e+000

B= 2.2408e-004

C= -2.3508e-002

Rho= 0.9997611

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/1)
0.0	0.000	-0.019	-0.019	18	0.000	18					-
0.2	0.200	0.185	-0.015	929	0.0 %	929					-8
2.5	2.500	2.554	0.054	11503	0.0 %	11503					2
5.0	5.000	5.076	0.076	22755	0.0 %	22755					2
7.5	7.500	7.340	-0.160	32861	0.0 %	32861					-2
10.0	10.000	10.065	0.065	45021	0.0 %	45021					1

LB118767 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 21 Feb 2022 12:55:07					
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7471B	PPB	21 Feb 2022 15:11:37	S	Std
0.2 - 1	50.2	929	-	0.2000	7471B	PPB	21 Feb 2022 15:13:54	S	Std
2.5 - 1	52.5	11503	-	2.5000	7471B	PPB	21 Feb 2022 15:16:12	S	Std
5.0 - 1	55	22755	-	5.0000	7471B	PPB	21 Feb 2022 15:18:29	S	Std
7.5 - 1	57.5	32861	-	7.5000	7471B	PPB	21 Feb 2022 15:20:46	S	Std
10.0 - 1	59.5	45021	-	10.0000	7471B	PPB	21 Feb 2022 15:28:19	S	Std
ICV99 - 1	ICV99	17289	3.8507	-	7471B	PPB	21 Feb 2022 15:31:06	U	SMPL
ICB99 - 1	ICB99	-29	-0.0300	-	7471B	PPB	21 Feb 2022 15:33:21	U	SMPL
CCV24 - 1	CCV24	21708	4.8409	-	7471B	PPB	21 Feb 2022 15:35:37	U	SMPL
CCB24 - 1	CCB24	-53	-0.0354	-	7471B	PPB	21 Feb 2022 15:37:53	U	SMPL
CRA - 1	CRA	889	0.1757	-	7471B	PPB	21 Feb 2022 15:40:10	U	SMPL
HighStd - 1	HighStd	44900	10.0379	-	7471B	PPB	21 Feb 2022 15:42:25	U	SMPL
ChkStd - 1	ChkStd	30474	6.8052	-	7471B	PPB	21 Feb 2022 15:44:41	U	SMPL
PB142818BL - 1	PBS	-34	-0.0311	-	7471B	PPB	21 Feb 2022 15:47:00	U	SMPL
PB142818BS - 1	LCSS	20035	4.4660	-	7471B	PPB	21 Feb 2022 15:49:19	U	SMPL
N1589-02 - 1	350	2425	0.5199	-	7471B	PPB	21 Feb 2022 15:51:36	U	SMPL
N1594-01 - 1	RO-CONCRETE-COMP	597	0.1103	-	7471B	PPB	21 Feb 2022 15:53:52	U	SMPL
N1598-01 - 1	GS-1	179	0.0166	-	7471B	PPB	21 Feb 2022 15:56:09	U	SMPL
N1598-01DUP - 1	GS-1DUP	174	0.0155	-	7471B	PPB	21 Feb 2022 15:58:27	U	SMPL
N1598-01MS - 1	GS-1MS	21989	4.9039	-	7471B	PPB	21 Feb 2022 16:00:42	U	SMPL
CCV25 - 1	CCV25	21591	4.8147	-	7471B	PPB	21 Feb 2022 16:02:58	U	SMPL
CCB25 - 1	CCB25	-35	-0.0314	-	7471B	PPB	21 Feb 2022 16:05:15	U	SMPL
N1598-01MSD - 1	GS-1MSD	20782	4.6334	-	7471B	PPB	21 Feb 2022 16:10:09	U	SMPL
N1598-02 - 1	GS-2	56	-0.0110	-	7471B	PPB	21 Feb 2022 16:12:24	U	SMPL
N1598-03 - 1	GS-3	266	0.0361	-	7471B	PPB	21 Feb 2022 16:14:40	U	SMPL
N1598-04 - 1	GS-4	639	0.1197	-	7471B	PPB	21 Feb 2022 16:16:56	U	SMPL
N1598-05 - 1	GS-5	231	0.0283	-	7471B	PPB	21 Feb 2022 16:19:12	U	SMPL
N1598-06 - 1	GS-6	463	0.0802	-	7471B	PPB	21 Feb 2022 16:21:29	U	SMPL
N1598-07 - 1	GS-7	3582	0.7792	-	7471B	PPB	21 Feb 2022 16:23:45	U	SMPL
N1598-08 - 1	GS-8	123	0.0041	-	7471B	PPB	21 Feb 2022 16:26:02	U	SMPL
PB142819BL - 1	PBS	8	-0.0217	-	7471B	PPB	21 Feb 2022 16:28:20	U	SMPL
PB142819BS - 1	LCSS	21469	4.7874	-	7471B	PPB	21 Feb 2022 16:30:35	U	SMPL
CCV26 - 1	CCV26	21028	4.6885	-	7471B	PPB	21 Feb 2022 16:32:52	U	SMPL
CCB26 - 1	CCB26	-56	-0.0361	-	7471B	PPB	21 Feb 2022 16:35:07	U	SMPL
N1602-01 - 1	20220075-BERKELEY-2-COMP	853	0.1676	-	7471B	PPB	21 Feb 2022 16:37:25	U	SMPL
N1602-02 - 1	20220076-BERKELEY-3-COMP	2174	0.4637	-	7471B	PPB	21 Feb 2022 16:39:40	U	SMPL
N1602-03 - 1	20220077A-BERKELEY-4A-COMP	10869	2.4121	-	7471B	PPB	21 Feb 2022 16:41:56	U	SMPL
N1602-04 - 1	20220077B-BERKELEY-4B-COMP	2069	0.4401	-	7471B	PPB	21 Feb 2022 16:44:13	U	SMPL
N1602-05 - 1	20220078-BERKELEY-6-COMP	17079	3.8036	-	7471B	PPB	21 Feb 2022 16:46:31	U	SMPL
N1602-06 - 1	20220082-AMENDED-BERKELEY-2-COMP	385	0.0628	-	7471B	PPB	21 Feb 2022 16:48:48	U	SMPL
N1602-07 - 1	20220083-AMENDED-BERKELEY-3-COMP	2610	0.5614	-	7471B	PPB	21 Feb 2022 16:51:08	U	SMPL
N1602-08 - 1	20220084A-AMENDED-BERKELEY-8A-COMP	9968	2.2102	-	7471B	PPB	21 Feb 2022 16:53:26	U	SMPL
N1602-09 - 1	20220084B-AMENDED-BERKELEY-4B-COMP	1914	0.4054	-	7471B	PPB	21 Feb 2022 16:55:44	U	SMPL
N1602-10 - 1	20220086-AMENDED-BERKELEY-6-COMP	16176	3.6013	-	7471B	PPB	21 Feb 2022 16:58:00	U	SMPL
CCV27 - 1	CCV27	20567	4.5852	-	7471B	PPB	21 Feb 2022 17:00:16	U	SMPL
CCB27 - 1	CCB27	-28	-0.0298	-	7471B	PPB	21 Feb 2022 17:02:32	U	SMPL
N1610-01 - 1	TR-05-021822	1887	0.3993	-	7471B	PPB	21 Feb 2022 17:09:38	U	SMPL
N1610-01DUP - 1	TR-05-021822DUP	1876	0.3969	-	7471B	PPB	21 Feb 2022 17:11:54	U	SMPL
N1610-01MS - 1	TR-05-021822MS	22928	5.1143	-	7471B	PPB	21 Feb 2022 17:14:09	U	SMPL
N1610-01MSD - 1	TR-05-021822MSD	23161	5.1665	-	7471B	PPB	21 Feb 2022 17:16:26	U	SMPL
N1611-04 - 1	SOIL-COMP	8571	1.8971	-	7471B	PPB	21 Feb 2022 17:18:41	U	SMPL
N1598-01LX5 - 1		25	-0.0179	-	7471B	PPB	21 Feb 2022 17:20:58	U	SMPL
N1598-01A - 1		18726	4.1727	-	7471B	PPB	21 Feb 2022 17:23:15	U	SMPL
N1610-01LX5 - 1		355	0.0560	-	7471B	PPB	21 Feb 2022 17:25:32	U	SMPL
N1610-01A - 1		18455	4.1120	-	7471B	PPB	21 Feb 2022 17:27:49	U	SMPL
CCV28 - 1	CCV28	21297	4.7488	-	7471B	PPB	21 Feb 2022 17:30:06	U	SMPL
CCB28 - 1	CCB28	-31	-0.0305	-	7471B	PPB	21 Feb 2022 17:32:22	U	SMPL