

LB118776


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39082



A= 0.0000e+000

B= 2.5272e-004

C= -2.4455e-002

Rho= 0.9996340

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.016	-0.016	32	0.000	32					-2
0.05	0.050	0.049	-0.001	290	0.0 %	290	0				-16
0.2	0.200	0.169	-0.031	766	0.0 %	766					-3
2.5	2.500	2.421	-0.079	9678	0.0 %	9678					-4
5.0	5.000	5.181	0.181	20596	0.0 %	20596					-1
7.5	7.500	7.594	0.094	30145	0.0 %	30145					-1
10.0	10.000	9.852	-0.148	39082	0.0 %	39082					-1

LB118776 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Feb 2022 12:51:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	245.1	PPB	22 Feb 2022 13:34:20	S	Std
0.05 - 1	50.05	290	-	0.0500	245.1	PPB	22 Feb 2022 13:36:37	S	Std
0.2 - 1	50.2	766	-	0.2000	245.1	PPB	22 Feb 2022 13:38:56	S	Std
2.5 - 1	52.5	9678	-	2.5000	245.1	PPB	22 Feb 2022 13:41:14	S	Std
5.0 - 1	55	20596	-	5.0000	245.1	PPB	22 Feb 2022 13:43:31	S	Std
7.5 - 1	57.5	30145	-	7.5000	245.1	PPB	22 Feb 2022 13:45:48	S	Std
10.0 - 1	510	39082	-	10.0000	245.1	PPB	22 Feb 2022 13:58:40	S	Std
ICV02 - 1	ICV02	15175	3.8106	-245.1		PPB	22 Feb 2022 14:10:18	U	SMPL
ICB02 - 1	ICB02	-32	-0.0325	-245.1		PPB	22 Feb 2022 14:12:33	U	SMPL
CCV32 - 1	CCV32	19336	4.8622	-245.1		PPB	22 Feb 2022 14:17:23	U	SMPL
CCB32 - 1	CCB32	-37	-0.0338	-245.1		PPB	22 Feb 2022 14:19:38	U	SMPL
CRA - 1	CRA	255	0.0400	-245.1		PPB	22 Feb 2022 14:21:54	U	SMPL
HighStd - 1	HighStd	38305	9.6561	-245.1		PPB	22 Feb 2022 14:24:09	U	SMPL
ChkStd - 1	ChkStd	25371	6.3874	-245.1		PPB	22 Feb 2022 14:26:24	U	SMPL
PB142849BL - 1	PBW	-59	-0.0394	-245.1		PPB	22 Feb 2022 14:28:42	U	SMPL
N1603-02 - 1	TI-604	60	-0.0093	-245.1		PPB	22 Feb 2022 14:33:17	U	SMPL
N1603-02DUP - 1	TI-604DUP	88	-0.0022	-245.1		PPB	22 Feb 2022 14:35:33	U	SMPL
N1603-02MS - 1	TI-604MS	13394	3.3605	-245.1		PPB	22 Feb 2022 14:40:44	U	SMPL
N1603-02MSD - 1	TI-604MSD	13335	3.3456	-245.1		PPB	22 Feb 2022 14:43:01	U	SMPL
N1604-02 - 1	NCQ-632	140	0.0109	-245.1		PPB	22 Feb 2022 14:45:17	U	SMPL
CCV33 - 1	CCV33	19208	4.8299	-245.1		PPB	22 Feb 2022 14:47:32	U	SMPL
CCB33 - 1	CCB33	-44	-0.0356	-245.1		PPB	22 Feb 2022 14:49:48	U	SMPL
PB142849BS - 1	LCSW	16585	4.1670	-245.1		PPB	22 Feb 2022 14:52:03	U	SMPL
N1603-02LX5 - 1		-27	-0.0313	-245.1		PPB	22 Feb 2022 14:54:19	U	SMPL
N1603-02A - 1		15270	3.8346	-245.1		PPB	22 Feb 2022 14:56:35	U	SMPL
CCV34 - 1	CCV34	18813	4.7300	-245.1		PPB	22 Feb 2022 14:58:51	U	SMPL
CCB34 - 1	CCB34	-38	-0.0341	-245.1		PPB	22 Feb 2022 15:01:07	U	SMPL