



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118778

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	02/22/2022	11:25
2	N1611-07		1	10	100.00	>212.0	>212.0	02/22/2022	11:40
3	N1611-07DUP		1	12	100.00	>212.0	>212.0	02/22/2022	12:05
4	N1611-08		1	15	100.00	>212.0	>212.0	02/22/2022	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP22222

WorkList ID : 157047

Department : Wet-Chemistry

Date : 02-22-2022 11:04:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2022	Water	N1611-07	Flash Point	Cool 4 deg C	PSEG03	M11	1070	02/18/2022	1010B
02/22/2022	Water	N1611-08	Flash Point	Cool 4 deg C	PSEG03	M11	WATER-COMP	02/18/2022	1010B

Date/Time 02/22/22 11:05
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 02/22/22 12:00
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)