



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118780

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/22/2022	13:25
2	CAL2	1	Water	NA	NA	20.3	7.00	02/22/2022	13:26
3	CAL3	1	Water	NA	NA	20.4	10.04	02/22/2022	13:27
4	ICV	1	Water	NA	NA	20.3	6.99	02/22/2022	13:30
5	CCV1	1	Water	NA	NA	20.4	2.01	02/22/2022	13:32
6	N1494-04	1	Solid	20.03	20	20.1	8.03	02/22/2022	13:33
7	N1494-04DUP	1	Solid	20.04	20	20.4	8.04	02/22/2022	13:35
8	CCV2	1	Water	NA	NA	20.4	12.03	02/22/2022	13:36

UB 118180

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1494 WorkList ID : 157055 Department : Wet-Chemistry Date : 02-22-2022 13:15:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/09/2022	Solid	N1494-04	pH	Cool 4 deg C	TULL02	M11	ASTORIA-YARD-PIT-10	02/09/2022	9045D

Date/Time 02/22/2022 13:20
Raw Sample Received by: JP (WCI)
Raw Sample Relinquished by: JP (WCI)

Date/Time 02/22/2022 14:00
Raw Sample Received by: JD (CSM)
Raw Sample Relinquished by: JP (WCI)