



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118782

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/22/2022	14:30
2	CAL2	1	Water	NA	NA	20.4	7.00	02/22/2022	14:31
3	CAL3	1	Water	NA	NA	20.4	10.02	02/22/2022	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	02/22/2022	14:35
5	CCV1	1	Water	NA	NA	20.4	2.01	02/22/2022	14:36
6	N1628-02	1	Solid	20.03	20	20.6	11.56	02/22/2022	14:38
7	N1628-02DUP	1	Solid	20.04	20	20.7	11.57	02/22/2022	14:39
8	CCV2	1	Water	NA	NA	20.5	12.03	02/22/2022	14:40

VB 118782

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n1635 S WorkList ID : 157059 Department : Wet-Chemistry Date : 02-22-2022 14:21:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
02/22/2022	Solid	N1628-02	Corrosivity	✓	Cool 4 deg C	SKAN06	N31	BP-3	02/22/2022 9045D

Date/Time 02/22/22 14:25
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CR SM

Date/Time 02/22/22 16:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JP (WC)