



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 2/23/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 02/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 02/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB118784

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1494-04	ASTORIA-YARD-PIT-10	1	1.12	8.73	9.85	8.86	88.7	
N1628-01	BP-3	2	1.14	8.52	9.66	8.28	83.8	
N1630-01	01	3	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1630-02	02	4	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1630-03	03	5	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1630-04	04	6	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1630-05	05	7	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N1633-01	20220099-MR-2-(DISCRETE)	8	1.15	8.50	9.65	7.15	70.6	sediment sample
N1633-02	20220104-MR-7-(DISCRETE)	9	1.15	8.83	9.98	7.02	66.5	sediment sample
N1633-03	20220117-COMPOISTE-C	10	1.18	8.63	9.81	7.04	67.9	sediment sample
N1634-01	20071	11	1.16	8.63	9.79	6.77	65.0	
N1634-02	20071-EPH-2	12	1.12	8.70	9.82	6.39	60.6	
N1635-01	RO-124055	13	1.15	8.80	9.95	4.62	39.4	GEL MATRIX
N1635-03	RO-124042	14	1.14	8.77	9.91	7.35	70.8	GEL MATRIX
N1635-05	RO-124019	15	1.13	8.84	9.97	7.99	77.6	GEL MATRIX
N1636-01	HR-02-022222	16	1.15	8.83	9.98	8.94	88.2	
N1636-02	HR-02-022222-EPH-2	17	1.18	8.52	9.7	8.65	87.7	
N1636-02D	HR-02-022222-EPH-2DUP	18	1.19	8.53	9.72	8.67	87.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$