


LB118785

7471B

INSTRUMENT ID: CVL

μ Abs.:

44714



Linear

A= 0.0000e+000

B= 2.2302e-004

C= 1.8082e-002

Rho= 0.9999871

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.018	0.018	1	0.000	1					1/1)
0.2	0.200	0.199	-0.001	809	0.0 %	809					1
2.5	2.500	2.465	-0.035	10971	0.0 %	10971					1
5.0	5.000	5.011	0.011	22387	0.0 %	22387					0
7.5	7.500	7.518	0.018	33628	0.0 %	33628					0
10.0	10.000	9.990	-0.010	44714	0.0 %	44714					0

LB118785 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Feb 2022 15:11:14							
Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type		
0.0 - 1	50	1	-	0.00007471B	PPB		22 Feb 2022 15:11:54	S	Std		
0.2 - 1	50.2	809	-	0.20007471B	PPB		22 Feb 2022 15:14:11	S	Std		
2.5 - 1	52.5	10971	-	2.50007471B	PPB		22 Feb 2022 15:16:29	S	Std		
5.0 - 1	55	22387	-	5.00007471B	PPB		22 Feb 2022 15:18:46	S	Std		
7.5 - 1	57.5	33628	-	7.50007471B	PPB		22 Feb 2022 15:21:04	S	Std		
10.0 - 1	510	44714	-	10.00007471B	PPB		22 Feb 2022 15:25:56	S	Std		
ICV03 - 1	ICV03	18074	4.0489	-7471B	PPB		22 Feb 2022 15:31:08	U	SMPL		
ICB03 - 1	ICB03	-56	0.0056	-7471B	PPB		22 Feb 2022 15:33:23	U	SMPL		
CCV35 - 1	CCV35	22871	5.1187	-7471B	PPB		22 Feb 2022 15:51:51	U	SMPL		
CCB35 - 1	CCB35	-65	0.0036	-7471B	PPB		22 Feb 2022 15:54:06	U	SMPL		
CRA - 1	CRA	896	0.2179	-7471B	PPB		22 Feb 2022 15:56:22	U	SMPL		
HighStd - 1	HighStd	42670	9.5342	-7471B	PPB		22 Feb 2022 15:58:37	U	SMPL		
ChkStd - 1	ChkStd	31283	6.9947	-7471B	PPB		22 Feb 2022 16:00:52	U	SMPL		
PB142865BL - 1	PBS	-60	0.0047	-7471B	PPB		22 Feb 2022 16:03:10	U	SMPL		
PB142865BS - 1	LCSS	18302	4.0997	-7471B	PPB		22 Feb 2022 16:05:27	U	SMPL		
N1626-01 - 1	CC-A1	1328	0.3142	-7471B	PPB		22 Feb 2022 16:07:43	U	SMPL		
N1626-02 - 1	CC-B1	7370	1.6617	-7471B	PPB		22 Feb 2022 16:10:01	U	SMPL		
N1626-02DUP - 1	CC-B1DUP	7634	1.7206	-7471B	PPB		22 Feb 2022 16:12:18	U	SMPL		
N1626-02MS - 1	CC-B1MS	25393	5.6811	-7471B	PPB		22 Feb 2022 16:14:35	U	SMPL		
N1626-02MSD - 1	CC-B1MSD	26114	5.8419	-7471B	PPB		22 Feb 2022 16:16:51	U	SMPL		
CCV36 - 1	CCV36	23608	5.2830	-7471B	PPB		22 Feb 2022 16:22:09	U	SMPL		
CCB36 - 1	CCB36	-51	0.0067	-7471B	PPB		22 Feb 2022 16:24:24	U	SMPL		
N1626-03 - 1	CC-A2	9273	2.0861	-7471B	PPB		22 Feb 2022 16:26:40	U	SMPL		
N1626-04 - 1	CC-B2	16920	3.7915	-7471B	PPB		22 Feb 2022 16:28:55	U	SMPL		
N1626-05 - 1	CC-A3	1639	0.3836	-7471B	PPB		22 Feb 2022 16:31:13	U	SMPL		
N1626-06 - 1	CC-B3	1139	0.2721	-7471B	PPB		22 Feb 2022 16:33:30	U	SMPL		
N1626-07 - 1	CC-A4	739	0.1829	-7471B	PPB		22 Feb 2022 16:35:46	U	SMPL		
N1626-08 - 1	CC-B4	566	0.1443	-7471B	PPB		22 Feb 2022 16:38:02	U	SMPL		
N1626-09 - 1	CC-A5	7182	1.6198	-7471B	PPB		22 Feb 2022 16:40:19	U	SMPL		
N1626-10 - 1	CC-B5	3534	0.8062	-7471B	PPB		22 Feb 2022 16:42:38	U	SMPL		
N1626-11 - 1	CC-A6	989	0.2386	-7471B	PPB		22 Feb 2022 16:44:56	U	SMPL		
N1626-12 - 1	CC-B6	1526	0.3584	-7471B	PPB		22 Feb 2022 16:47:12	U	SMPL		
CCV37 - 1	CCV37	22590	5.0560	-7471B	PPB		22 Feb 2022 16:49:28	U	SMPL		
CCB37 - 1	CCB37	-48	0.0074	-7471B	PPB		22 Feb 2022 16:51:45	U	SMPL		
N1626-13 - 1	CC-A7	1448	0.3410	-7471B	PPB		22 Feb 2022 16:54:01	U	SMPL		
N1626-14 - 1	CC-B7	2461	0.5669	-7471B	PPB		22 Feb 2022 16:56:17	U	SMPL		
N1626-15 - 1	CC-A8	429	0.1138	-7471B	PPB		22 Feb 2022 16:58:33	U	SMPL		
N1626-16 - 1	CC-B8	8589	1.9336	-7471B	PPB		22 Feb 2022 17:00:49	U	SMPL		
N1633-01 - 1	20220099-MR-2-(DISCRETE)	4552	1.0332	-7471B	PPB		22 Feb 2022 17:03:05	U	SMPL		
N1633-02 - 1	20220104-MR-7-(DISCRETE)	747	0.1847	-7471B	PPB		22 Feb 2022 17:05:21	U	SMPL		
N1633-03 - 1	20220117-COMPOISTE-C	5395	1.2213	-7471B	PPB		22 Feb 2022 17:07:38	U	SMPL		
N1626-02LX5 - 1		1603	0.3756	-7471B	PPB		22 Feb 2022 17:09:56	U	SMPL		
N1626-02A - 1		26268	5.8763	-7471B	PPB		22 Feb 2022 17:12:14	U	SMPL		
CCV38 - 1	CCV38	23725	5.3091	-7471B	PPB		22 Feb 2022 17:14:30	U	SMPL		
CCB38 - 1	CCB38	-44	0.0083	-7471B	PPB		22 Feb 2022 17:16:47	U	SMPL		