



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118792

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/23/2022	11:10
2	CAL2	1	Water	NA	NA	20.4	7.00	02/23/2022	11:11
3	CAL3	1	Water	NA	NA	20.4	10.03	02/23/2022	11:12
4	ICV	1	Water	NA	NA	20.5	7.01	02/23/2022	11:15
5	CCV1	1	Water	NA	NA	20.5	2.01	02/23/2022	11:19
6	N1639-01	1	Solid	20.03	20	21.3	5.61	02/23/2022	11:20
7	N1639-01DUP	1	Solid	20.04	20	21.4	5.62	02/23/2022	11:21
8	CCV2	1	Water	NA	NA	20.5	12.03	02/23/2022	11:25

213 118792

WORKLIST(Hardcopy Internal Chain)

Date : 02-23-2022 10:58:14

WorkList ID : 157081 Department : Wet-Chemistry

WorkList Name : PH N1639 S

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/22/2022	Solid	N1639-01	pH	Cool 4 deg C	NOBI03	M11	WP-SOIL-1-FINAL	02/22/2022	9045D

Date/Time 02/23/22 11:00
Raw Sample Received by: [Signature] 50162C1
Raw Sample Relinquished by: [Signature]

Date/Time 02/23/22 12:30
Raw Sample Received by: [Signature] 50162C1
Raw Sample Relinquished by: [Signature]