



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118800

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.02	02/23/2022	16:35
2	CAL2	1	Water	NA	NA	20.8	7.00	02/23/2022	16:37
3	CAL3	1	Water	NA	NA	20.6	10.02	02/23/2022	16:38
4	ICV	1	Water	NA	NA	20.8	6.99	02/23/2022	16:40
5	CCV1	1	Water	NA	NA	20.6	2.00	02/23/2022	16:42
6	N1638-01	1	Water	NA	NA	20.1	6.50	02/23/2022	16:45
7	N1638-01DUP	1	Water	NA	NA	20.1	6.50	02/23/2022	16:46
8	N1638-02	1	Water	NA	NA	20.4	6.45	02/23/2022	16:48
9	N1642-01	1	Water	NA	NA	20.4	5.76	02/23/2022	16:50
10	N1642-02	1	Water	NA	NA	20.3	6.03	02/23/2022	16:51
11	N1642-03	1	Water	NA	NA	20.4	6.06	02/23/2022	16:52
12	N1642-04	1	Water	NA	NA	20.6	5.83	02/23/2022	16:54
13	N1642-05	1	Water	NA	NA	20.7	5.94	02/23/2022	16:55
14	N1642-06	1	Water	NA	NA	20.6	6.07	02/23/2022	16:57
15	N1642-08	1	Water	NA	NA	20.7	6.06	02/23/2022	16:58
16	CCV2	1	Water	NA	NA	20.7	12.02	02/23/2022	17:00
17	N1647-01	1	Water	NA	NA	20.6	5.93	02/23/2022	17:02
18	N1647-02	1	Water	NA	NA	20.4	5.98	02/23/2022	17:03
19	N1647-03	1	Water	NA	NA	20.4	6.44	02/23/2022	17:05
20	CCV3	1	Water	NA	NA	20.6	2.00	02/23/2022	17:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH22322 WorkList ID : 157092 Department : Wet-Chemistry Date : 02-23-2022 16:25:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
02/23/2022	Water	N1638-01	pH	Cool 4 deg C	PSEG03	M11	30960	02/23/2022	9040C
02/23/2022	Water	N1638-02	pH	Cool 4 deg C	PSEG03	M11	40473	02/23/2022	9040C
02/23/2022	Water	N1642-01	pH	Cool 4 deg C	PSEG03	M11	31453	02/23/2022	9040C
02/23/2022	Water	N1642-02	pH	Cool 4 deg C	PSEG03	M11	31454	02/23/2022	9040C
02/23/2022	Water	N1642-03	pH	Cool 4 deg C	PSEG03	M11	31455	02/23/2022	9040C
02/23/2022	Water	N1642-04	pH	Cool 4 deg C	PSEG03	M11	31456	02/23/2022	9040C
02/23/2022	Water	N1642-05	pH	Cool 4 deg C	PSEG03	M11	31457	02/23/2022	9040C
02/23/2022	Water	N1642-06	pH	Cool 4 deg C	PSEG03	M11	31458	02/23/2022	9040C
02/23/2022	Water	N1642-08	pH	Cool 4 deg C	PSEG03	M11	31409	02/23/2022	9040C
02/23/2022	Water	N1647-01	pH	Cool 4 deg C	PSEG03	P11	31382	02/23/2022	9040C
02/23/2022	Water	N1647-02	pH	Cool 4 deg C	PSEG03	P11	31415	02/23/2022	9040C
02/23/2022	Water	N1647-03	pH	Cool 4 deg C	PSEG03	P11	31416	02/23/2022	9040C

Date/Time 02/23/22 16:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 02/23/22 17:10
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)