



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118801

Slope : 98.8

BalanceID: WC-SC-4

pH Meter ID : WC PH MERER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.02	02/23/2022	16:35
2	CAL2	1	Water	NA	NA	20.8	7.00	02/23/2022	16:37
3	CAL3	1	Water	NA	NA	20.6	10.02	02/23/2022	16:38
4	ICV	1	Water	NA	NA	20.8	6.99	02/23/2022	16:40
5	CCV1	1	Water	NA	NA	20.5	2.00	02/23/2022	17:08
6	N1650-02	1	Solid	20	20	20.9	8.92	02/23/2022	17:10
7	N1650-02DUP	1	Solid	20	20	20.9	8.91	02/23/2022	17:11
8	N1650-04	1	Solid	20	20	20.7	8.53	02/23/2022	17:12
9	CCV2	1	Water	NA	NA	20.6	12.02	02/23/2022	17:13

W3118801

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n1650 WorkList ID : 157091 Department : Wet-Chemistry Date : 02-23-2022 16:12:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
02/23/2022	Solid	N1650-02	Corrosivity	Cool 4 deg C	ENTA05	N22	SW-SURFCIAL-SOIL-(SW-SG)	02/23/2022	9045D
02/23/2022	Solid	N1650-04	Corrosivity	Cool 4 deg C	ENTA05	N22	NW-SURFCIAL-SOIL-(NW-SG)	02/23/2022	9045D

Date/Time	02/23/22	16:10
Raw Sample Received by:	JP (WC)	
Raw Sample Relinquished by:	JP (WC)	

Page 1 of 1