

LA118803


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

37626



A= 0.0000e+000

B= 2.6442e-004

C= 6.7338e-002

Rho= 0.9998101

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.073	0.073	21	0.000	21					21
0.2	0.200	0.241	0.041	655	0.0 %	655					-5
2.5	2.500	2.387	-0.113	8773	0.0 %	8773					-2
5.0	5.000	4.918	-0.082	18343	0.0 %	18343					1
7.5	7.500	7.565	0.065	28356	0.0 %	28356					0
10.0	10.000	10.017	0.017	37626	0.0 %	37626					

LB118803 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 24 Feb 2022 09:33:22					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	24 Feb 2022 09:37:17	S	Std
0.2 - 1	50.2	655	-	0.2000	7470A	PPB	24 Feb 2022 09:39:34	S	Std
2.5 - 1	52.5	8773	-	2.5000	7470A	PPB	24 Feb 2022 09:41:52	S	Std
5.0 - 1	55	18343	-	5.0000	7470A	PPB	24 Feb 2022 09:44:09	S	Std
7.5 - 1	57.5	28356	-	7.5000	7470A	PPB	24 Feb 2022 09:46:25	S	Std
10.0 - 1	510	37626	-	10.0000	7470A	PPB	24 Feb 2022 10:01:29	S	Std
ICV05 - 1	ICV05	15610	4.1950	-7470A	PPB	24 Feb 2022 10:19:25	U	SMPL	
ICB05 - 1	ICB05	-18	0.0626	-7470A	PPB	24 Feb 2022 10:21:40	U	SMPL	
CCV42 - 1	CCV42	20427	5.4687	-7470A	PPB	24 Feb 2022 10:23:55	U	SMPL	
CCB42 - 1	CCB42	-30	0.0594	-7470A	PPB	24 Feb 2022 10:26:11	U	SMPL	
CRA - 1	CRA	675	0.2458	-7470A	PPB	24 Feb 2022 10:28:27	U	SMPL	
HighStd - 1	HighStd	36941	9.8354	-7470A	PPB	24 Feb 2022 10:30:42	U	SMPL	
ChkStd - 1	ChkStd	28223	7.5301	-7470A	PPB	24 Feb 2022 10:32:57	U	SMPL	
PB142878BL - 1	PBW	111	0.0967	-7470A	PPB	24 Feb 2022 10:35:14	U	SMPL	
N1628-02 - 1	BP-3	480	0.1943	-7470A	PPB	24 Feb 2022 10:39:47	U	SMPL	
N1628-02DUP - 1	BP-3DUP	525	0.2062	-7470A	PPB	24 Feb 2022 10:42:03	U	SMPL	
N1628-02MS - 1	BP-3MS	15562	4.1823	-7470A	PPB	24 Feb 2022 10:44:21	U	SMPL	
N1628-02MSD - 1	BP-3MSD	15877	4.2656	-7470A	PPB	24 Feb 2022 10:46:38	U	SMPL	
N1635-02 - 1	RO-124055	-23	0.0613	-7470A	PPB	24 Feb 2022 10:48:53	U	SMPL	
CCV43 - 1	CCV43	17995	4.8256	-7470A	PPB	24 Feb 2022 10:51:09	U	SMPL	
CCB43 - 1	CCB43	-33	0.0586	-7470A	PPB	24 Feb 2022 10:53:24	U	SMPL	
N1635-04 - 1	RO-124042	21	0.0729	-7470A	PPB	24 Feb 2022 10:55:40	U	SMPL	
N1635-06 - 1	RO-124019	15	0.0713	-7470A	PPB	24 Feb 2022 10:57:55	U	SMPL	
PB142879BL - 1	PBW	13	0.0708	-7470A	PPB	24 Feb 2022 11:00:11	U	SMPL	
PB142879BS - 1	LCSW	16317	4.3819	-7470A	PPB	24 Feb 2022 11:02:27	U	SMPL	
N1632-01 - 1	20220090-BERKELEY-2-ELUTIRATE	33	0.0761	-7470A	PPB	24 Feb 2022 11:04:43	U	SMPL	
N1632-01DUP - 1	20220090-BERKELEY-2-ELUTIRATEDUP	52	0.0811	-7470A	PPB	24 Feb 2022 11:07:00	U	SMPL	
N1632-01MS - 1	20220090-BERKELEY-2-ELUTIRATEMS	17849	4.7870	-7470A	PPB	24 Feb 2022 11:09:16	U	SMPL	
N1632-01MSD - 1	20220090-BERKELEY-2-ELUTIRATEMSD	17705	4.7489	-7470A	PPB	24 Feb 2022 11:11:33	U	SMPL	
N1632-02 - 1	20220092B-BERKELEY-4B-ELUTRIATE	58	0.0827	-7470A	PPB	24 Feb 2022 11:13:50	U	SMPL	
N1632-03 - 1	20220094-BERKELEY-6-ELUTRIATE	75	0.0872	-7470A	PPB	24 Feb 2022 11:16:06	U	SMPL	
CCV44 - 1	CCV44	19496	5.2225	-7470A	PPB	24 Feb 2022 11:22:06	U	SMPL	
CCB44 - 1	CCB44	-65	0.0502	-7470A	PPB	24 Feb 2022 11:24:22	U	SMPL	
PB142867TB - 1	PB142867TB	24	0.0737	-7470A	PPB	24 Feb 2022 11:38:18	U	SMPL	
PB142878BS - 1	LCSW	15328	4.1204	-7470A	PPB	24 Feb 2022 11:40:33	U	SMPL	
N1628-02LX5 - 1		3	0.0681	-7470A	PPB	24 Feb 2022 11:42:49	U	SMPL	
N1628-02A - 1		15676	4.2124	-7470A	PPB	24 Feb 2022 11:45:05	U	SMPL	
N1632-01LX5 - 1		-7	0.0655	-7470A	PPB	24 Feb 2022 11:47:22	U	SMPL	
N1632-01A - 1		18009	4.8293	-7470A	PPB	24 Feb 2022 11:49:39	U	SMPL	
CCV45 - 1	CCV45	20120	5.3875	-7470A	PPB	24 Feb 2022 11:51:56	U	SMPL	
CCB45 - 1	CCB45	-28	0.0599	-7470A	PPB	24 Feb 2022 11:54:13	U	SMPL	